

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091018\
 Data File : VN051101.D
 Acq On : 11 Sep 2018 1:36
 Operator : MD\SY
 Sample : J4777-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-05-090718-C

Quant Time: Sep 11 05:31:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	623386	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	971749	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	838683	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282096	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	373380	53.82	ug/l	0.00
Spiked Amount						
			Recovery	=	107.64%	
35) Dibromofluoromethane	7.59	113	367322	50.31	ug/l	0.00
Spiked Amount						
			Recovery	=	100.62%	
50) Toluene-d8	10.09	98	1327899	47.92	ug/l	0.00
Spiked Amount						
			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	12.40	95	366630	40.00	ug/l	0.00
Spiked Amount						
			Recovery	=	80.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	14472	2.66	ug/l	96
18) Methyl Acetate	4.33	43	9314	1.68	ug/l #	67
20) Methylene Chloride	4.55	84	5691963	742.77	ug/l	100
30) Chloroform	7.38	83	14012	1.10	ug/l	91
43) Isopropyl Acetate	8.17	43	219938	20.17	ug/l #	85
68) m/p-Xylenes	11.62	106	13567	1.03	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	32320	1.72	ug/l	98
95) Naphthalene	15.13	128	8966	6.10	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091018\
Data File : VN051101.D
Acq On : 11 Sep 2018 1:36
Operator : MD\SY
Sample : J4777-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-05-090718-C

Quant Time: Sep 11 05:31:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

