

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057805.D
 Acq On : 9 Sep 2019 16:54
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
Client Sampled :
 I.BLK

Manual Integrations
APPROVED
 MMDadoda
 9/10/2019 4:30:23 PM

Quant Time: Sep 10 03:42:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	505542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	828965	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	701191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	285800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	339201	33.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	66.26%	
35) Dibromofluoromethane	7.57	113	241656	31.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	63.16%	
50) Toluene-d8	10.09	98	1009544	33.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	67.98%	
62) 4-Bromofluorobenzene	12.40	95	322056	28.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	56.90%	
Target Compounds						
17) Carbon Disulfide	4.03	76	7574	1.060	ug/l #	84
88) 1,4-Dichlorobenzene	13.37	146	8773m	0.852	ug/l	
93) 1,2,4-Trichlorobenzene	14.91	180	14701	2.218	ug/l	98
95) Naphthalene	15.13	128	66640	2.490	ug/l	98
96) 1,2,3-Trichlorobenzene	15.30	180	19421	4.288	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057805.D
 Acq On : 9 Sep 2019 16:54
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampled :
 I.BLK

Manual Integrations
APPROVED
 MMDadoda
 9/10/2019 4:30:23 PM

Quant Time: Sep 10 03:42:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

