

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057825.D
 Acq On : 10 Sep 2019 00:38
 Operator : JC/SP
 Sample : K4603-33
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-12-F3-(0)

Quant Time: Sep 13 04:42:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	289396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	494724	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	457150	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	227785	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	333968	56.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.96%	
35) Dibromofluoromethane	7.58	113	223140	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	10.08	98	898142	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	12.40	95	295528	43.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.48%	
Target Compounds						
16) Acetone	3.81	43	10440	5.433	ug/l	90
20) Methylene Chloride	4.53	84	12244	2.170	ug/l	96
43) Isopropyl Acetate	8.15	43	83422	8.368	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057825.D
Acq On : 10 Sep 2019 00:38
Operator : JC/SP
Sample : K4603-33
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-12-F3-(0)

Quant Time: Sep 13 04:42:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

