

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057845.D
 Acq On : 10 Sep 2019 8:42
 Operator : JC/SP
 Sample : PB122910TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB122910TB

Manual Integrations
 APPROVED

MMDadoda
 9/13/2019 5:16:29 AM

Quant Time: Sep 10 11:05:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	154896	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	326238	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	261576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	104124	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287517	91.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	183.28%	
35) Dibromofluoromethane	7.58	113	184534	61.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.58%	
50) Toluene-d8	10.09	98	506698	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
62) 4-Bromofluorobenzene	12.40	95	156633	35.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.32%	
Target Compounds						
16) Acetone	3.81	43	9461	9.199	ug/l #	86
20) Methylene Chloride	4.52	84	6053m	1.908	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057845.D
Acq On : 10 Sep 2019 8:42
Operator : JC/SP
Sample : PB122910TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB122910TB

Manual Integrations
APPROVED

MMDadoda
9/13/2019 5:16:29 AM

Quant Time: Sep 10 11:05:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

