

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057849.D
 Acq On : 10 Sep 2019 10:10
 Operator : JC/SP
 Sample : K4781-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW800-8.0-13.0-090719

Quant Time: Sep 10 11:12:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	287171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	483975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	435651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	307730	52.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.82%	
35) Dibromofluoromethane	7.58	113	206303	46.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.36%	
50) Toluene-d8	10.09	98	839085	48.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.76%	
62) 4-Bromofluorobenzene	12.40	95	271019	41.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.02%	
Target Compounds						
16) Acetone	3.81	43	22192	11.639	ug/l	98
30) Chloroform	7.36	83	11912	1.485	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057849.D
Acq On : 10 Sep 2019 10:10
Operator : JC/SP
Sample : K4781-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW800-8.0-13.0-090719

Quant Time: Sep 10 11:12:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

