

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057850.D
 Acq On : 10 Sep 2019 10:32
 Operator : JC/SP
 Sample : K4781-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW801-8.0-13.0-090719

Quant Time: Sep 10 11:14:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	290618	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	499209	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	447085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	232337	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	318943	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
35) Dibromofluoromethane	7.58	113	218050	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
50) Toluene-d8	10.09	98	888595	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	12.40	95	288610	42.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.68%	
Target Compounds						
16) Acetone	3.81	43	20720	10.738	ug/l	90
30) Chloroform	7.36	83	21443	2.642	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057850.D
Acq On : 10 Sep 2019 10:32
Operator : JC/SP
Sample : K4781-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW801-8.0-13.0-090719

Quant Time: Sep 10 11:14:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

