

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057851.D
 Acq On : 10 Sep 2019 10:56
 Operator : JC/SP
 Sample : K4622-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-B2-(0)

Manual Integrations
 APPROVED

MMDadoda
 9/13/2019 8:16:41 AM

Quant Time: Sep 13 04:59:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	297616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	510098	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	471467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	246566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	327493	54.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.66%	
35) Dibromofluoromethane	7.58	113	221873	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
50) Toluene-d8	10.08	98	935241	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	12.40	95	303209	43.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.06%	
Target Compounds						
16) Acetone	3.80	43	11040	5.587	ug/l	98
43) Isopropyl Acetate	8.16	43	100190m	9.747	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057851.D
Acq On : 10 Sep 2019 10:56
Operator : JC/SP
Sample : K4622-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-B2-(0)

Manual Integrations
APPROVED

MMDadoda
9/13/2019 8:16:41 AM

Quant Time: Sep 13 04:59:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

