

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057852.D
 Acq On : 10 Sep 2019 11:18
 Operator : JC/SP
 Sample : K4622-33
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-B3-(0)

Manual Integrations
 APPROVED

MMDadoda
 9/13/2019 8:16:42 AM

Quant Time: Sep 13 04:59:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	294397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	520971	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	486228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	243773	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	334726	56.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.28%	
35) Dibromofluoromethane	7.58	113	228786	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
50) Toluene-d8	10.09	98	945423	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	12.40	95	306995	43.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.30%	
Target Compounds						
16) Acetone	3.80	43	9264	4.739	ug/l	99
43) Isopropyl Acetate	8.16	43	97895m	9.325	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057852.D
Acq On : 10 Sep 2019 11:18
Operator : JC/SP
Sample : K4622-33
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
PAR-11-B3-(0)

Manual Integrations
APPROVED

MMDadoda
9/13/2019 8:16:42 AM

Quant Time: Sep 13 04:59:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

