

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057856.D
 Acq On : 10 Sep 2019 12:45
 Operator : JC/SP
 Sample : K4622-36
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-B2-(4)

Quant Time: Sep 10 13:49:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	239674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	431251	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	385866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	189685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	302481	62.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.62%	
35) Dibromofluoromethane	7.57	113	203519	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	10.08	98	752204	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	12.40	95	246939	41.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.86%	
Target Compounds						
16) Acetone	3.80	43	13906	8.738	ug/l	95
20) Methylene Chloride	4.53	84	13720	3.381	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057856.D
Acq On : 10 Sep 2019 12:45
Operator : JC/SP
Sample : K4622-36
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 62 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-B2-(4)

Quant Time: Sep 10 13:49:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

