

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091019\
 Data File : VN057858.D
 Acq On : 10 Sep 2019 13:29
 Operator : JC/SP
 Sample : K4603-29
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-E4-(3.11)

Quant Time: Sep 10 13:57:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	273108	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	470875	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	427328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	217489	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	298523	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
35) Dibromofluoromethane	7.58	113	203448	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
50) Toluene-d8	10.09	98	818504	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
62) 4-Bromofluorobenzene	12.40	95	266604	41.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.92%	
Target Compounds						
16) Acetone	3.80	43	9679	5.338	ug/l	91
20) Methylene Chloride	4.53	84	12801	2.540	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091019\
Data File : VN057858.D
Acq On : 10 Sep 2019 13:29
Operator : JC/SP
Sample : K4603-29
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-E4-(3.11)

Quant Time: Sep 10 13:57:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

