

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091020\
 Data File : VN063296.D
 Acq On : 10 Sep 2020 15:36
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN091020

Quant Time: Sep 21 18:44:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 21 18:39:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.480	0.456	5.0	99	0.00
3 P	Chloromethane	0.719	0.654	9.0	100	0.00
4 C	Vinyl Chloride	0.679	0.666	1.9#	98	0.00
5 T	Bromomethane	0.474	0.462	2.5	101	0.00
6 T	Chloroethane	0.444	0.418	5.9	96	0.00
7 T	Trichlorofluoromethane	1.051	0.983	6.5	96	0.00
8 T	Diethyl Ether	0.375	0.358	4.5	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.549	0.537	2.2	100	0.00
10 T	Methyl Iodide	0.712	0.682	4.2	92	0.00
11 T	Tert butyl alcohol	0.104	0.103	1.0	99	0.00
12 CM	1,1-Dichloroethene	0.548	0.526	4.0#	99	0.00
13 T	Acrolein	0.054	0.056	-3.7	112	0.00
14 T	Allyl chloride	1.125	1.068	5.1	98	0.00
15 T	Acrylonitrile	0.284	0.278	2.1	97	0.00
16 T	Acetone	0.271	0.257	5.2	101	0.00
17 T	Carbon Disulfide	1.698	1.552	8.6	99	0.00
18 T	Methyl Acetate	0.682	0.643	5.7	98	0.00
19 T	Methyl tert-butyl Ether	1.958	1.903	2.8	98	0.00
20 T	Methylene Chloride	0.691	0.624	9.7	100	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.575	6.2	98	0.00
22 T	Diisopropyl ether	2.214	2.170	2.0	98	0.00
23 T	Vinyl Acetate	1.742	1.808	-3.8	98	0.00
24 P	1,1-Dichloroethane	1.199	1.168	2.6	99	0.00
25 T	2-Butanone	0.401	0.394	1.7	98	0.00
26 T	2,2-Dichloropropane	1.051	1.022	2.8	100	0.00
27 T	cis-1,2-Dichloroethene	0.715	0.689	3.6	100	0.00
28 T	Bromochloromethane	0.581	0.556	4.3	101	0.00
29 T	Tetrahydrofuran	0.261	0.256	1.9	96	0.00
30 C	Chloroform	1.200	1.168	2.7#	98	0.00
31 T	Cyclohexane	1.102	1.043	5.4	98	0.00
32 T	1,1,1-Trichloroethane	1.046	1.023	2.2	99	0.00
33 S	1,2-Dichloroethane-d4	0.786	0.748	4.8	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.323	0.318	1.5	96	0.00
36 T	1,1-Dichloropropene	0.514	0.512	0.4	99	0.00
37 T	Ethyl Acetate	0.506	0.489	3.4	97	0.00
38 T	Carbon Tetrachloride	0.506	0.511	-1.0	99	0.00
39 T	Methylcyclohexane	0.561	0.566	-0.9	100	0.00
40 TM	Benzene	1.468	1.457	0.7	98	0.00
41 T	Methacrylonitrile	0.229	0.222	3.1	92	-0.02
42 TM	1,2-Dichloroethane	0.557	0.551	1.1	96	0.00
43 T	Isopropyl Acetate	0.856	0.859	-0.4	97	0.00
44 TM	Trichloroethene	0.384	0.374	2.6	95	0.00
45 C	1,2-Dichloropropane	0.403	0.396	1.7#	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091020\
 Data File : VN063296.D
 Acq On : 10 Sep 2020 15:36
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 ICVVN091020

Quant Time: Sep 21 18:44:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 21 18:39:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.261	0.261	0.0	100	0.00
47 T	Bromodichloromethane	0.544	0.544	0.0	99	0.00
48 T	Methyl methacrylate	0.409	0.406	0.7	96	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	99	0.00
50 S	Toluene-d8	1.236	1.225	0.9	96	0.00
51 T	4-Methyl-2-Pentanone	0.472	0.487	-3.2	97	0.00
52 CM	Toluene	0.896	0.906	-1.1#	99	0.00
53 T	t-1,3-Dichloropropene	0.593	0.609	-2.7	99	0.00
54 T	cis-1,3-Dichloropropene	0.639	0.647	-1.3	97	0.00
55 T	1,1,2-Trichloroethane	0.351	0.346	1.4	97	0.00
56 T	Ethyl methacrylate	0.534	0.553	-3.6	98	0.00
57 T	1,3-Dichloropropane	0.614	0.617	-0.5	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.286	0.275	3.8	99	0.00
59 T	2-Hexanone	0.348	0.360	-3.4	97	0.00
60 T	Dibromochloromethane	0.386	0.398	-3.1	99	0.00
61 T	1,2-Dibromoethane	0.366	0.366	0.0	99	0.00
62 S	4-Bromofluorobenzene	0.457	0.459	-0.4	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.410	0.408	0.5	100	0.00
65 PM	Chlorobenzene	1.076	1.046	2.8	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.384	2.0	98	0.00
67 C	Ethyl Benzene	1.932	1.962	-1.6#	99	0.00
68 T	m/p-Xylenes	0.715	0.717	-0.3	100	0.00
69 T	o-Xylene	0.686	0.689	-0.4	99	0.00
70 T	Styrene	1.140	1.179	-3.4	100	0.00
71 P	Bromoform	0.271	0.275	-1.5	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	4.033	3.907	3.1	100	0.00
74 T	N-amyl acetate	1.656	1.740	-5.1	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.155	1.071	7.3	98	0.00
76 T	1,2,3-Trichloropropane	1.103	1.066	3.4	101	0.00
77 T	Bromobenzene	0.927	0.873	5.8	99	0.00
78 T	n-propylbenzene	4.582	4.585	-0.1	100	0.00
79 T	2-Chlorotoluene	2.817	2.716	3.6	100	0.00
80 T	1,3,5-Trimethylbenzene	3.349	3.325	0.7	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.411	0.399	2.9	98	0.00
82 T	4-Chlorotoluene	2.953	2.889	2.2	103	0.00
83 T	tert-Butylbenzene	2.840	2.723	4.1	100	0.00
84 T	1,2,4-Trimethylbenzene	3.375	3.346	0.9	100	0.00
85 T	sec-Butylbenzene	3.613	3.648	-1.0	102	0.00
86 T	p-Isopropyltoluene	3.313	3.352	-1.2	103	0.00
87 T	1,3-Dichlorobenzene	1.695	1.638	3.4	101	0.00
88 T	1,4-Dichlorobenzene	1.759	1.662	5.5	102	0.00
89 T	n-Butylbenzene	2.990	3.097	-3.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091020\
Data File : VN063296.D
Acq On : 10 Sep 2020 15:36
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN091020

Quant Time: Sep 21 18:44:01 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 21 18:39:24 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.545	0.573	-5.1	101	0.00
91 T	1,2-Dichlorobenzene	1.675	1.612	3.8	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.252	4.2	96	0.00
93 T	1,2,4-Trichlorobenzene	0.980	0.950	3.1	101	0.00
94 T	Hexachlorobutadiene	0.417	0.360	13.7	100	0.00
95 T	Naphthalene	3.179	2.923	8.1	97	0.00
96 T	1,2,3-Trichlorobenzene	0.957	0.885	7.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6