

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083749.D
 Acq On : 10 Sep 2024 12:43
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 11 01:58:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 01:38:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	259980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	44799	19.406	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	38.820%#
35) Dibromofluoromethane	8.159	113	33344	20.034	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	40.060%#
50) Toluene-d8	10.565	98	117482	20.418	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	40.840%#
62) 4-Bromofluorobenzene	12.847	95	52334	22.046	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	44.100%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	37820	21.271	ug/l	99
3) Chloromethane	2.359	50	42184	22.207	ug/l	95
4) Vinyl Chloride	2.512	62	42371	21.741	ug/l	93
5) Bromomethane	2.901	94	25120	22.449	ug/l	97
6) Chloroethane	3.071	64	29793	22.035	ug/l	92
7) Trichlorofluoromethane	3.471	101	75588	22.849	ug/l	98
8) Diethyl Ether	3.959	74	25651	23.131	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.353	101	41166	23.477	ug/l	98
10) Methyl Iodide	4.577	142	50200	22.511	ug/l	97
11) Tert butyl alcohol	5.542	59	35218	113.923	ug/l	98
12) 1,1-Dichloroethene	4.330	96	38356	23.162	ug/l	97
13) Acrolein	4.171	56	28707	94.116	ug/l	99
14) Allyl chloride	5.012	41	67821	22.985	ug/l	97
15) Acrylonitrile	5.718	53	96660	114.975	ug/l	100
16) Acetone	4.430	43	121335	131.110	ug/l	94
17) Carbon Disulfide	4.700	76	99836	21.553	ug/l	99
18) Methyl Acetate	5.018	43	45786	22.733	ug/l	98
19) Methyl tert-butyl Ether	5.794	73	139123	23.217	ug/l	97
20) Methylene Chloride	5.265	84	43066	22.740	ug/l	94
21) trans-1,2-Dichloroethene	5.777	96	39218	22.099	ug/l	90
22) Diisopropyl ether	6.665	45	141158	23.239	ug/l	96
23) Vinyl Acetate	6.594	43	528904	118.904	ug/l	99
24) 1,1-Dichloroethane	6.565	63	79312	22.664	ug/l	99
25) 2-Butanone	7.483	43	150475	123.710	ug/l	100
26) 2,2-Dichloropropane	7.483	77	77312	23.990	ug/l	98
27) cis-1,2-Dichloroethene	7.483	96	48657	23.096	ug/l	98
28) Bromochloromethane	7.812	49	27780	18.851	ug/l	99
29) Tetrahydrofuran	7.841	42	83865	116.606	ug/l	99
30) Chloroform	7.959	83	87460	23.542	ug/l	90
31) Cyclohexane	8.253	56	73847	22.133	ug/l	95
32) 1,1,1-Trichloroethane	8.165	97	79170	23.176	ug/l	93
36) 1,1-Dichloropropene	8.365	75	57819	23.081	ug/l	99
37) Ethyl Acetate	7.559	43	54774	24.088	ug/l	97
38) Carbon Tetrachloride	8.359	117	66860	23.526	ug/l	99
39) Methylcyclohexane	9.600	83	68785	24.123	ug/l	97
40) Benzene	8.600	78	180938	23.997	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083749.D
 Acq On : 10 Sep 2024 12:43
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 11 01:58:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 01:38:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	31140	23.154	ug/l	98
42) 1,2-Dichloroethane	8.665	62	68226	24.216	ug/l	99
43) Isopropyl Acetate	8.688	43	97966	23.775	ug/l	98
44) Trichloroethene	9.353	130	40782	23.356	ug/l	96
45) 1,2-Dichloropropane	9.618	63	44560	24.451	ug/l	98
46) Dibromomethane	9.706	93	30954	24.250	ug/l	97
47) Bromodichloromethane	9.882	83	66147	23.945	ug/l #	94
48) Methyl methacrylate	9.677	41	46788	23.933	ug/l	98
49) 1,4-Dioxane	9.694	88	14984	465.914	ug/l #	96
51) 4-Methyl-2-Pentanone	10.441	43	270783	120.721	ug/l	97
52) Toluene	10.629	92	110278	23.911	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	65148	23.266	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	70696	24.035	ug/l	94
55) 1,1,2-Trichloroethane	11.018	97	40247	23.375	ug/l	96
56) Ethyl methacrylate	10.871	69	65086	24.569	ug/l	96
57) 1,3-Dichloropropane	11.159	76	70916	23.964	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	97384	102.024	ug/l	98
59) 2-Hexanone	11.194	43	210382	126.795	ug/l	98
60) Dibromochloromethane	11.359	129	45716	23.725	ug/l	99
61) 1,2-Dibromoethane	11.471	107	40643	23.669	ug/l	98
64) Tetrachloroethene	11.106	164	34851	22.688	ug/l	93
65) Chlorobenzene	11.888	112	116122	22.620	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	41153	23.124	ug/l	98
67) Ethyl Benzene	11.965	91	210713	23.106	ug/l	97
68) m/p-Xylenes	12.071	106	158230	46.734	ug/l	100
69) o-Xylene	12.400	106	76822	23.438	ug/l	99
70) Styrene	12.412	104	125129	23.262	ug/l	98
71) Bromoform	12.576	173	27805	23.185	ug/l #	99
73) Isopropylbenzene	12.694	105	198534	21.379	ug/l	100
74) N-amyl acetate	12.494	43	83761	21.141	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	55934	20.500	ug/l	97
76) 1,2,3-Trichloropropane	12.994	75	75078	30.999	ug/l	63
77) Bromobenzene	12.982	156	44670	19.984	ug/l	100
78) n-propylbenzene	13.035	91	229309	21.223	ug/l	99
79) 2-Chlorotoluene	13.123	91	147252	21.370	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	167379	21.379	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	18748	20.274	ug/l #	80
82) 4-Chlorotoluene	13.223	91	143958	20.818	ug/l	99
83) tert-Butylbenzene	13.441	119	147261	21.499	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	167470	21.294	ug/l	100
85) sec-Butylbenzene	13.618	105	197117	21.699	ug/l	98
86) p-Isopropyltoluene	13.729	119	160758	21.110	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	84190	19.957	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	82492	19.311	ug/l	99
89) n-Butylbenzene	14.059	91	140204	20.506	ug/l	99
90) Hexachloroethane	14.335	117	31578	20.629	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	83557	20.523	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	11364	20.893	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	42008	19.228	ug/l	98
94) Hexachlorobutadiene	15.500	225	19047	20.528	ug/l	99
95) Naphthalene	15.641	128	128956	19.401	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	40335	19.317	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083749.D
Acq On : 10 Sep 2024 12:43
Operator : JC\MD
Sample : VSTDIC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Sep 11 01:58:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 01:38:19 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 11 01:58:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 01:38:19 2024
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.224 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:168 Resp: 153102

Ion Ratio Lower Upper

168 100

99 68.6 54.2 81.2



#2

Dichlorodifluoromethane

Concen: 21.271 ug/l

RT: 2.124 min Scan# 29

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 85 Resp: 37820

Ion Ratio Lower Upper

85 100

87 31.8 15.7 47.0





#3

Chloromethane

Concen: 22.207 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 50 Resp: 42184

Ion Ratio Lower Upper

50 100

52 31.1 27.2 40.8



#4

Vinyl Chloride

Concen: 21.741 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 62 Resp: 42371

Ion Ratio Lower Upper

62 100

64 27.2 24.7 37.1









#5
Bromomethane
Concen: 22.449 ug/l
RT: 2.901 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 25120
Ion Ratio Lower Upper
94 100
96 91.9 76.0 114.0



#6
Chloroethane
Concen: 22.035 ug/l
RT: 3.071 min Scan# 190
Delta R.T. -0.012 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 64 Resp: 29793
Ion Ratio Lower Upper
64 100
66 34.3 23.9 35.9









#7
Trichlorofluoromethane
Concen: 22.849 ug/l
RT: 3.471 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 75588
Ion Ratio Lower Upper
101 100
103 62.9 51.4 77.0



#8
Diethyl Ether
Concen: 23.131 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 74 Resp: 25651
Ion Ratio Lower Upper
74 100
45 97.8 46.7 140.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 23.477 ug/l
RT: 4.353 min Scan# 408
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 41166
Ion Ratio Lower Upper
101 100
85 48.3 36.2 54.4
151 71.8 57.7 86.5



#10
Methyl Iodide
Concen: 22.511 ug/l
RT: 4.577 min Scan# 446
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:142 Resp: 50200
Ion Ratio Lower Upper
142 100
127 45.8 37.5 56.3
141 16.4 11.0 16.4









#11
Tert butyl alcohol
Concen: 113.923 ug/l
RT: 5.542 min Scan# 610
Delta R.T. 0.012 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 35218
Ion Ratio Lower Upper
59 100
57 10.3 8.7 13.1



#12
1,1-Dichloroethene
Concen: 23.162 ug/l
RT: 4.330 min Scan# 404
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 96 Resp: 38356
Ion Ratio Lower Upper
96 100
61 175.6 143.4 215.0
98 63.4 52.7 79.1





#13

Acrolein

Concen: 94.116 ug/l

RT: 4.171 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 28707

Ion Ratio Lower Upper

56 100

55 72.0 58.2 87.2



#14

Allyl chloride

Concen: 22.985 ug/l

RT: 5.012 min Scan# 520

Delta R.T. -0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion: 41 Resp: 67821


Ion Ratio Lower Upper

41 100

39 77.1 59.3 88.9

76 34.4 27.7 41.5









#15
Acrylonitrile
Concen: 114.975 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53	Resp: 96660
Ion Ratio	Lower Upper
53 100	
52 82.1	66.1 99.1
51 36.9	29.6 44.4



#16
Acetone
Concen: 131.110 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 43	Resp: 121335
Ion Ratio	Lower Upper
43 100	
58 26.5	23.7 35.5









#17
Carbon Disulfide
Concen: 21.553 ug/l
RT: 4.700 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 99836
Ion Ratio Lower Upper
76 100
78 8.5 7.1 10.7



#18
Methyl Acetate
Concen: 22.733 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 43 Resp: 45786
Ion Ratio Lower Upper
43 100
74 22.9 19.0 28.4





#19

Methyl tert-butyl Ether

Concen: 23.217 ug/l

RT: 5.794 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 73 Resp: 139123

Ion Ratio Lower Upper

73 100

57 20.0 17.0 25.6



#20

Methylene Chloride

Concen: 22.740 ug/l

RT: 5.265 min Scan# 563

Delta R.T. -0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Tgt Ion: 84 Resp: 43066




Ion Ratio Lower Upper


84 100

49 126.3 93.4 140.0

51 35.3 29.2 43.8

86 60.0 46.9 70.3









#21
trans-1,2-Dichloroethene
Concen: 22.099 ug/l
RT: 5.777 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 39218
Ion Ratio Lower Upper
96 100
61 146.5 106.0 159.0
98 61.3 52.2 78.2



#22
Diisopropyl ether
Concen: 23.239 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 45 Resp: 141158
Ion Ratio Lower Upper
45 100
43 56.0 43.4 65.0
87 24.8 22.6 33.8
59 10.8 9.1 13.7









#23
 Vinyl Acetate
 Concen: 118.904 ug/l
 RT: 6.594 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 528904
 Ion Ratio Lower Upper
 43 100
 86 10.6 8.7 13.1



#24
 1,1-Dichloroethane
 Concen: 22.664 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.006 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43

Tgt Ion: 63 Resp: 79312
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.5 10.4
 100 4.6 2.5 7.5









#25
2-Butanone
Concen: 123.710 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 150475
Ion Ratio Lower Upper
43 100
72 26.2 20.8 31.2



#26
2,2-Dichloropropane
Concen: 23.990 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 77 Resp: 77312
Ion Ratio Lower Upper
77 100
97 21.3 10.2 30.4









#27
 cis-1,2-Dichloroethene
 Concen: 23.096 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43

Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion: 96 Resp: 48657




Ion	Ratio	Lower	Upper
96	100		
61	156.3	0.0	306.2
98	64.8	0.0	128.6




#28
 Bromochloromethane
 Concen: 18.851 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43

Tgt Ion: 49 Resp: 27780

Ion	Ratio	Lower	Upper
49	100		
129	0.9	0.0	3.4
130	75.7	60.2	90.2





#29

Tetrahydrofuran

Concen: 116.606 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min


Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	42	Resp:	83865
Ion Ratio	Lower	Upper	
42	100		
72	44.3	35.9	53.9
71	41.6	33.9	50.9



#30

Chloroform

Concen: 23.542 ug/l




RT: 7.959 min Scan# 1021


Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Tgt Ion:	83	Resp:	87460
Ion Ratio	Lower	Upper	
83	100		
85	57.7	52.8	79.2





#31

Cyclohexane

Concen: 22.133 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min


Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 56 Resp: 73847

Ion Ratio Lower Upper

56 100

69 33.2 26.9 40.3

84 92.3 69.2 103.8



#32

1,1,1-Trichloroethane

Concen: 23.176 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion: 97 Resp: 79170


Ion Ratio Lower Upper

97 100

99 55.9 52.3 78.5

61 47.5 38.2 57.2









#33
1,2-Dichloroethane-d4
Concen: 19.406 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 65 Resp: 44799
Ion Ratio Lower Upper
65 100
67 49.9 0.0 100.8



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:114 Resp: 259980
Ion Ratio Lower Upper
114 100
63 22.7 0.0 43.4
88 17.0 0.0 32.6





#35

Dibromofluoromethane

Concen: 20.034 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. -0.006 min

Lab File: VN083749.D





Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:113	Resp:	33344
Ion Ratio	Lower	Upper
113	100	
111	101.7	82.6 124.0
192	16.0	14.3 21.5



#36

1,1-Dichloropropene

Concen: 23.081 ug/l




RT: 8.365 min Scan# 1090


Delta R.T. -0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43


Tgt Ion: 75	Resp:	57819
Ion Ratio	Lower	Upper
75	100	
110	33.2	16.7 50.1
77	30.7	25.2 37.8








#37
Ethyl Acetate
Concen: 24.088 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	43	Resp:	54774
Ion	Ratio	Lower	Upper
43	100		
61	12.2	11.1	16.7
70	10.4	8.8	13.2



#38
Carbon Tetrachloride
Concen: 23.526 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:	117	Resp:	66860
Ion	Ratio	Lower	Upper
117	100		
119	96.8	76.1	114.1
121	30.3	24.1	36.1





#39

Methylcyclohexane

Concen: 24.123 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 68785

Ion Ratio Lower Upper

83 100

55 75.7 62.2 93.4

98 50.6 37.6 56.4



#40

Benzene

Concen: 23.997 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 78 Resp: 180938

Ion Ratio Lower Upper

78 100

77 23.3 19.0 28.6









#41
Methacrylonitrile
Concen: 23.154 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	41	Resp:	31140
Ion	Ratio	Lower	Upper
41	100		
39	58.8	45.4	68.2
67	69.1	55.9	83.9
52	32.9	24.1	36.1



#42
1,2-Dichloroethane
Concen: 24.216 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:	62	Resp:	68226
Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	16.6









#43
Isopropyl Acetate
Concen: 23.775 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
43	100		
61	25.3	19.4	29.0
87	12.6	9.8	14.6



#44
Trichloroethene
Concen: 23.356 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion	Ratio	Lower	Upper
130	100		
95	103.9	0.0	200.4









#45
1,2-Dichloropropane
Concen: 24.451 ug/l
RT: 9.618 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 44560
Ion Ratio Lower Upper
63 100
65 31.3 25.9 38.9



#46
Dibromomethane
Concen: 24.250 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 93 Resp: 30954
Ion Ratio Lower Upper
93 100
95 81.0 65.3 97.9
174 84.2 71.0 106.4









#47
Bromodichloromethane
Concen: 23.945 ug/l
RT: 9.882 min Scan# 1349
Delta R.T. -0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 66147
Ion Ratio Lower Upper
83 100
85 60.4 51.8 77.8
127 6.2 6.6 9.8#



#48
Methyl methacrylate
Concen: 23.933 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43


Tgt Ion: 41 Resp: 46788
Ion Ratio Lower Upper
41 100
69 79.8 64.3 96.5
39 64.6 48.9 73.3








#49
1,4-Dioxane
Concen: 465.914 ug/l
RT: 9.694 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 88 Resp: 14984
Ion Ratio Lower Upper
88 100
43 0.0 0.0 0.0
58 74.8 57.1 85.7



#50
Toluene-d8
Concen: 20.418 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 98 Resp: 117482
Ion Ratio Lower Upper
98 100
100 63.4 52.0 78.0





#51

4-Methyl-2-Pentanone

Concen: 120.721 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 270783

Ion Ratio Lower Upper

43 100

58 37.6 31.6 47.4



#52

Toluene

Concen: 23.911 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 92 Resp: 110278

Ion Ratio Lower Upper

92 100

91 170.5 134.7 202.1









#53
t-1,3-Dichloropropene
Concen: 23.266 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 65148
Ion Ratio Lower Upper
75 100
77 30.2 25.8 38.6



#54
cis-1,3-Dichloropropene
Concen: 24.035 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 75 Resp: 70696
Ion Ratio Lower Upper
75 100
77 30.8 25.0 37.6
39 59.5 42.6 64.0





#55

1,1,2-Trichloroethane

Concen: 23.375 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 97 Resp: 40247





Ion Ratio Lower Upper

97 100

83 81.8 67.5 101.3

85 57.8 44.2 66.2

99 58.4 50.1 75.1



#56

Ethyl methacrylate

Concen: 24.569 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion: 69 Resp: 65086


Ion Ratio Lower Upper

69 100

41 69.9 53.9 80.9

39 44.6 32.6 48.8





#57

1,3-Dichloropropane

Concen: 23.964 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 70916

Ion Ratio Lower Upper

76 100

78 32.6 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 102.024 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 63 Resp: 97384

Ion Ratio Lower Upper

63 100

106 28.1 23.3 34.9





#59

2-Hexanone

Concen: 126.795 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 210382

Ion Ratio Lower Upper

43 100

58 53.1 27.2 81.5



#60

Dibromochloromethane

Concen: 23.725 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion:129 Resp: 45716

Ion Ratio Lower Upper

129 100

127 78.4 39.0 116.9





#61

1,2-Dibromoethane

Concen: 23.669 ug/l

RT: 11.471 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:107 Resp: 40643

Ion Ratio Lower Upper

107 100

109 92.5 75.6 113.4



#62

4-Bromofluorobenzene

Concen: 22.046 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion: 95 Resp: 52334


Ion Ratio Lower Upper

95 100

174 69.6 0.0 139.8

176 65.7 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:117 Resp: 230981

Ion Ratio Lower Upper

117 100

82 61.1 47.4 71.0

119 33.2 26.3 39.5



#64

Tetrachloroethene

Concen: 22.688 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Tgt Ion:164 Resp: 34851




Ion Ratio Lower Upper


164 100

166 118.1 103.7 155.5

129 96.4 79.9 119.9

131 91.7 78.4 117.6









#65
Chlorobenzene
Concen: 22.620 ug/l
RT: 11.888 min Scan# 116122
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 116122
Ion Ratio Lower Upper
112 100
114 30.5 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 23.124 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:131 Resp: 41153
Ion Ratio Lower Upper
131 100
133 98.4 47.5 142.6
119 62.7 31.6 94.8









#67
Ethyl Benzene
Concen: 23.106 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 210713
Ion Ratio Lower Upper
91 100
106 28.9 24.3 36.5



#68
m/p-Xylenes
Concen: 46.734 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:106 Resp: 158230
Ion Ratio Lower Upper
106 100
91 211.5 169.2 253.8





#69

o-Xylene

Concen: 23.438 ug/l

RT: 12.400 min Scan# 1775

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:106 Resp: 76822

Ion Ratio Lower Upper

106 100

91 226.6 112.9 338.6



#70

Styrene

Concen: 23.262 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion:104 Resp: 125129


Ion Ratio Lower Upper

104 100

78 55.7 43.3 64.9

103 55.6 43.4 65.2









#71
Bromoform
Concen: 23.185 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 27805
Ion Ratio Lower Upper
173 100
175 48.8 24.1 72.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion:152 Resp: 120031
Ion Ratio Lower Upper
152 100
115 64.1 32.2 96.6
150 162.3 0.0 348.6





#73

Isopropylbenzene

Concen: 21.379 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 198534

Ion Ratio Lower Upper

105 100

120 25.1 12.6 37.8



#74

N-ethyl acetate

Concen: 21.141 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Tgt Ion: 43 Resp: 83761




Ion Ratio Lower Upper


43 100

70 41.0 34.5 51.7

55 24.5 18.6 28.0

61 23.8 19.2 28.8









#75
1,1,2,2-Tetrachloroethane
Concen: 20.500 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 55934
Ion Ratio Lower Upper
83 100
131 10.2 5.2 15.6
85 66.3 31.9 95.9



#76
1,2,3-Trichloropropane
Concen: 30.999 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43

Tgt Ion: 75 Resp: 75078
Ion Ratio Lower Upper
75 100
77 135.7 95.2 285.6





#77

Bromobenzene

Concen: 19.984 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:156 Resp: 44670

Ion Ratio Lower Upper

156 100

77 228.1 114.3 343.1

158 97.6 48.9 146.7



#78

n-propylbenzene

Concen: 21.223 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 91 Resp: 229309

Ion Ratio Lower Upper

91 100

120 21.4 11.0 33.0





#79

2-Chlorotoluene

Concen: 21.370 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 91 Resp: 147252

Ion Ratio Lower Upper

91 100

126 30.3 15.4 46.2



#80

1,3,5-Trimethylbenzene

Concen: 21.379 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 105 Resp: 167379

Ion Ratio Lower Upper

105 100

120 46.6 23.4 70.3









#81
trans-1,4-Dichloro-2-butene
Concen: 20.274 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 18748
Ion Ratio Lower Upper
75 100
53 127.4 81.2 121.8#
89 49.9 35.7 53.5



#82
4-Chlorotoluene
Concen: 20.818 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43


Tgt Ion: 91 Resp: 143958
Ion Ratio Lower Upper
91 100
126 31.3 15.4 46.2








#83
tert-Butylbenzene
Concen: 21.499 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.006 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:119 Resp: 147261
Ion Ratio Lower Upper
119 100
91 71.1 34.5 103.6
134 25.3 12.8 38.4




#84
1,2,4-Trimethylbenzene
Concen: 21.294 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083749.D
Acq: 10 Sep 2024 12:43



Tgt Ion:105 Resp: 167470
Ion Ratio Lower Upper
105 100
120 43.8 21.8 65.4





#85

sec-Butylbenzene

Concen: 21.699 ug/l

RT: 13.618 min Scan# 1983

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 197117

Ion Ratio Lower Upper

105 100

134 18.6 9.7 29.1



#86

p-Isopropyltoluene

Concen: 21.110 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion:119 Resp: 160758


Ion Ratio Lower Upper

119 100

134 25.6 12.8 38.3

91 26.5 12.9 38.7





#87

1,3-Dichlorobenzene

Concen: 19.957 ug/l

RT: 13.735 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 84190

Ion Ratio Lower Upper

146 100

111 43.5 21.7 65.1

148 64.5 31.7 95.1



#88

1,4-Dichlorobenzene

Concen: 19.311 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43




Tgt Ion:146 Resp: 82492


Ion Ratio Lower Upper

146 100

111 44.3 21.3 64.0

148 64.3 32.0 96.2





#89

n-Butylbenzene

Concen: 20.506 ug/l

RT: 14.059 min Scan# 2057

Delta R.T. 0.006 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 91 Resp: 140204

Ion Ratio Lower Upper

91 100

92 52.4 26.7 80.1

134 24.1 12.2 36.4



#90

Hexachloroethane

Concen: 20.629 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.006 min

Lab File: VN083749.D




Acq: 10 Sep 2024 12:43


Tgt Ion: 117 Resp: 31578

Ion Ratio Lower Upper

117 100

201 65.0 32.8 98.3









#91
 1,2-Dichlorobenzene
 Concen: 20.523 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.000 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:146 Resp: 83557
 Ion Ratio Lower Upper
 146 100
 111 45.8 22.8 68.3
 148 62.8 32.1 96.5



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 20.893 ug/l
 RT: 14.723 min Scan# 2171
 Delta R.T. 0.006 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43

Tgt Ion: 75 Resp: 11364
 Ion Ratio Lower Upper
 75 100
 155 71.9 35.9 107.7
 157 89.8 47.5 142.5









#93
 1,2,4-Trichlorobenzene
 Concen: 19.228 ug/l
 RT: 15.388 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:180	Resp:	42008
Ion Ratio	Lower	Upper
180	100	
182	94.2	48.1 144.3
145	33.4	16.9 50.7



#94
 Hexachlorobutadiene
 Concen: 20.528 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: VN083749.D
 Acq: 10 Sep 2024 12:43

Tgt Ion:225	Resp:	19047
Ion Ratio	Lower	Upper
225	100	
223	63.2	31.9 95.5
227	62.5	32.0 96.2





#95

Naphthalene

Concen: 19.401 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.006 min


Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion:128 Resp: 128956

Ion Ratio Lower Upper

128 100

127 12.6 10.2 15.2

129 10.6 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 19.317 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.000 min

Lab File: VN083749.D

Acq: 10 Sep 2024 12:43

Tgt Ion:180 Resp: 40335

Ion Ratio Lower Upper

180 100

182 95.7 47.1 141.4

145 34.7 17.8 53.5

