

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083768.D
 Acq On : 11 Sep 2024 03:39
 Operator : JC\MD
 Sample : VN0910WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02

Quant Time: Sep 11 05:17:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 03:05:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	143387	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257368	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	227885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112878	52.208	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.420%			
35) Dibromofluoromethane	8.165	113	82589	50.126	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.260%			
50) Toluene-d8	10.565	98	289156	50.765	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.520%			
62) 4-Bromofluorobenzene	12.847	95	111431	47.417	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.840%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.118	85	31867	19.137	ug/l	100
3) Chloromethane	2.360	50	34655	19.244	ug/l	94
4) Vinyl Chloride	2.512	62	34675	18.998	ug/l	91
5) Bromomethane	2.901	94	20152	19.230	ug/l	93
6) Chloroethane	3.071	64	23261	18.370	ug/l	96
7) Trichlorofluoromethane	3.471	101	58582	18.908	ug/l	95
8) Diethyl Ether	3.959	74	19933	19.193	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.354	101	30512	18.580	ug/l	96
10) Methyl Iodide	4.571	142	39121	18.731	ug/l	96
11) Tert butyl alcohol	5.530	59	27662	95.544	ug/l	99
12) 1,1-Dichloroethene	4.318	96	28891	18.629	ug/l #	83
13) Acrolein	4.177	56	26041	91.160	ug/l	99
14) Allyl chloride	5.012	41	52108	18.856	ug/l	98
15) Acrylonitrile	5.712	53	77308	98.187	ug/l	99
16) Acetone	4.424	43	68043	78.506	ug/l	98
17) Carbon Disulfide	4.701	76	76667	17.672	ug/l	99
18) Methyl Acetate	5.024	43	38960	20.479	ug/l	99
19) Methyl tert-butyl Ether	5.795	73	107388	19.135	ug/l	99
20) Methylene Chloride	5.271	84	34011	19.176	ug/l	94
21) trans-1,2-Dichloroethene	5.777	96	30294	18.227	ug/l	95
22) Diisopropyl ether	6.671	45	112691	19.810	ug/l	96
23) Vinyl Acetate	6.600	43	397068	95.314	ug/l	98
24) 1,1-Dichloroethane	6.559	63	62720	19.137	ug/l	98
25) 2-Butanone	7.483	43	105108	92.267	ug/l	98
26) 2,2-Dichloropropane	7.483	77	42783	14.175	ug/l	98
27) cis-1,2-Dichloroethene	7.483	96	37380	18.946	ug/l	99
28) Bromochloromethane	7.812	49	28180	20.418	ug/l	94
29) Tetrahydrofuran	7.842	42	66007	97.994	ug/l	99
30) Chloroform	7.965	83	66187	19.023	ug/l	90
31) Cyclohexane	8.247	56	57697	18.464	ug/l	97
32) 1,1,1-Trichloroethane	8.159	97	62627	19.575	ug/l	98
36) 1,1-Dichloropropene	8.371	75	44245	17.841	ug/l	98
37) Ethyl Acetate	7.559	43	43951	19.525	ug/l	99
38) Carbon Tetrachloride	8.359	117	51584	18.335	ug/l	98
39) Methylcyclohexane	9.600	83	49877	17.670	ug/l	98
40) Benzene	8.600	78	141057	18.898	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083768.D
 Acq On : 11 Sep 2024 03:39
 Operator : JC\MD
 Sample : VN0910WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02

Quant Time: Sep 11 05:17:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 03:05:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	26053	19.568	ug/l	97
42) 1,2-Dichloroethane	8.671	62	53060	19.024	ug/l	100
43) Isopropyl Acetate	8.689	43	80290	19.683	ug/l	99
44) Trichloroethene	9.353	130	31364	18.145	ug/l	97
45) 1,2-Dichloropropane	9.618	63	34140	18.924	ug/l	95
46) Dibromomethane	9.706	93	24478	19.371	ug/l	98
47) Bromodichloromethane	9.888	83	51921	18.986	ug/l	100
48) Methyl methacrylate	9.683	41	35820	18.508	ug/l	96
49) 1,4-Dioxane	9.694	88	11512	361.588	ug/l #	98
51) 4-Methyl-2-Pentanone	10.447	43	213921	96.339	ug/l	98
52) Toluene	10.630	92	84819	18.577	ug/l	95
53) t-1,3-Dichloropropene	10.835	75	47380	17.092	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	52323	17.969	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	32129	18.849	ug/l	94
56) Ethyl methacrylate	10.877	69	49523	18.884	ug/l	96
57) 1,3-Dichloropropane	11.159	76	55493	18.943	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	87631	92.738	ug/l	99
59) 2-Hexanone	11.194	43	153851	93.665	ug/l	98
60) Dibromochloromethane	11.359	129	35833	18.785	ug/l	99
61) 1,2-Dibromoethane	11.471	107	31767	18.687	ug/l	98
64) Tetrachloroethene	11.106	164	27325	18.031	ug/l	97
65) Chlorobenzene	11.888	112	88414	17.456	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	31580	17.986	ug/l	97
67) Ethyl Benzene	11.959	91	159960	17.779	ug/l	96
68) m/p-Xylenes	12.071	106	117861	35.284	ug/l	99
69) o-Xylene	12.394	106	59388	18.365	ug/l	97
70) Styrene	12.412	104	96345	18.154	ug/l	99
71) Bromoform	12.576	173	21734	18.369	ug/l #	98
73) Isopropylbenzene	12.694	105	149372	17.792	ug/l	98
74) N-amyl acetate	12.494	43	64934	18.128	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	45224	18.333	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	56520	25.813	ug/l	67
77) Bromobenzene	12.976	156	34088	16.868	ug/l	96
78) n-propylbenzene	13.035	91	171132	17.519	ug/l	99
79) 2-Chlorotoluene	13.124	91	111010	17.820	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	128127	18.102	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	13477	16.120	ug/l	96
82) 4-Chlorotoluene	13.218	91	106816	17.086	ug/l	100
83) tert-Butylbenzene	13.435	119	111318	17.976	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	126783	17.831	ug/l	99
85) sec-Butylbenzene	13.612	105	145907	17.766	ug/l	98
86) p-Isopropyltoluene	13.729	119	118583	17.224	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	63010	16.521	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	62787	16.258	ug/l	99
89) n-Butylbenzene	14.053	91	96128	15.552	ug/l	100
90) Hexachloroethane	14.329	117	24745	17.880	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	63786	17.329	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.718	75	8779	17.853	ug/l	95
93) 1,2,4-Trichlorobenzene	15.388	180	29965	15.171	ug/l	99
94) Hexachlorobutadiene	15.500	225	13045	15.552	ug/l	96
95) Naphthalene	15.641	128	96573	16.071	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	29500	15.627	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083768.D
Acq On : 11 Sep 2024 03:39
Operator : JC\MD
Sample : VN0910WBS02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Quant Time: Sep 11 05:17:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 03:05:46 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02







Tgt Ion:168 Resp: 143387
Ion Ratio Lower Upper
168 100
99 64.0 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 19.137 ug/l
RT: 2.118 min Scan# 28
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 85 Resp: 31867
Ion Ratio Lower Upper
85 100
87 31.3 15.7 47.0





#3

Chloromethane

Concen: 19.244 ug/l

RT: 2.360 min Scan# 69

Delta R.T. 0.001 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion: 50 Resp: 34655

Ion Ratio Lower Upper

50 100

52 30.8 27.2 40.8



#4

Vinyl Chloride

Concen: 18.998 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN083768.D




Acq: 11 Sep 2024 03:39

Tgt Ion: 62 Resp: 34675

Ion Ratio Lower Upper

62 100

64 36.1 24.7 37.1









#5
Bromomethane
Concen: 19.230 ug/l
RT: 2.901 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion: 94 Resp: 20152
Ion Ratio Lower Upper
94 100
96 88.3 76.0 114.0



#6
Chloroethane
Concen: 18.370 ug/l
RT: 3.071 min Scan# 190
Delta R.T. -0.012 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 64 Resp: 23261
Ion Ratio Lower Upper
64 100
66 32.3 23.9 35.9








#7
 Trichlorofluoromethane
 Concen: 18.908 ug/l
 RT: 3.471 min Scan# 258
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02

Tgt Ion:101 Resp: 58582
 Ion Ratio Lower Upper
 101 100
 103 60.4 51.4 77.0



#8
 Diethyl Ether
 Concen: 19.193 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39

Tgt Ion: 74 Resp: 19933
 Ion Ratio Lower Upper
 74 100
 45 95.9 46.7 140.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.580 ug/l
RT: 4.354 min Scan# 408
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion:	101	Resp:	30512
Ion	Ratio	Lower	Upper
101	100		
85	49.5	36.2	54.4
151	74.7	57.7	86.5



#10
Methyl Iodide
Concen: 18.731 ug/l
RT: 4.571 min Scan# 445
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:	142	Resp:	39121
Ion	Ratio	Lower	Upper
142	100		
127	49.1	37.5	56.3
141	15.5	11.0	16.4









#11
Tert butyl alcohol
Concen: 95.544 ug/l
RT: 5.530 min Scan# 608
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 59 Resp: 27662
Ion Ratio Lower Upper
59 100
57 11.3 8.7 13.1



#12
1,1-Dichloroethene
Concen: 18.629 ug/l
RT: 4.318 min Scan# 402
Delta R.T. -0.012 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 96 Resp: 28891
Ion Ratio Lower Upper
96 100
61 156.0 143.4 215.0
98 51.1 52.7 79.1#





#13

Acrolein

Concen: 91.160 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion: 56 Resp: 26041

Ion Ratio Lower Upper

56 100

55 71.6 58.2 87.2



#14

Allyl chloride

Concen: 18.856 ug/l

RT: 5.012 min Scan# 520

Delta R.T. -0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion: 41 Resp: 52108

Ion Ratio Lower Upper

41 100

39 74.8 59.3 88.9

76 32.5 27.7 41.5









#15
Acrylonitrile
Concen: 98.187 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:	53	Resp:	77308
Ion	Ratio	Lower	Upper
53	100		
52	81.8	66.1	99.1
51	35.8	29.6	44.4



#16
Acetone
Concen: 78.506 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:	43	Resp:	68043
Ion	Ratio	Lower	Upper
43	100		
58	30.9	23.7	35.5









#17
Carbon Disulfide
Concen: 17.672 ug/l
RT: 4.701 min Scan# 467
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion: 76 Resp: 76667
Ion Ratio Lower Upper
76 100
78 9.1 7.1 10.7



#18
Methyl Acetate
Concen: 20.479 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 43 Resp: 38960
Ion Ratio Lower Upper
43 100
74 23.4 19.0 28.4





#19

Methyl tert-butyl Ether

Concen: 19.135 ug/l

RT: 5.795 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion: 73 Resp: 107388

Ion Ratio Lower Upper

73 100

57 20.8 17.0 25.6



#20

Methylene Chloride

Concen: 19.176 ug/l

RT: 5.271 min Scan# 564

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Tgt Ion: 84 Resp: 34011




Ion Ratio Lower Upper

84 100

49 120.7 93.4 140.0

51 38.7 29.2 43.8

86 67.3 46.9 70.3









#21
trans-1,2-Dichloroethene
Concen: 18.227 ug/l
RT: 5.777 min Scan# 610
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 96 Resp: 30294
Ion Ratio Lower Upper
96 100
61 139.8 106.0 159.0
98 62.1 52.2 78.2



#22
Diisopropyl ether
Concen: 19.810 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 45 Resp: 112691
Ion Ratio Lower Upper
45 100
43 56.3 43.4 65.0
87 24.9 22.6 33.8
59 12.0 9.1 13.7









#23
 Vinyl Acetate
 Concen: 95.314 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02

Tgt Ion: 43 Resp: 397068
 Ion Ratio Lower Upper
 43 100
 86 10.0 8.7 13.1



#24
 1,1-Dichloroethane
 Concen: 19.137 ug/l
 RT: 6.559 min Scan# 783
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39

Tgt Ion: 63 Resp: 62720
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.5 10.4
 100 4.0 2.5 7.5





#25

2-Butanone

Concen: 92.267 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion: 43 Resp: 105108

Ion Ratio Lower Upper

43 100

72 25.0 20.8 31.2



#26

2,2-Dichloropropane

Concen: 14.175 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN083768.D




Acq: 11 Sep 2024 03:39

Tgt Ion: 77 Resp: 42783

Ion Ratio Lower Upper

77 100

97 21.4 10.2 30.4








#27
 cis-1,2-Dichloroethene
 Concen: 18.946 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02


Tgt Ion: 96 Resp: 37380
 Ion Ratio Lower Upper
 96 100
 61 150.7 0.0 306.2
 98 64.3 0.0 128.6



#28
 Bromochloromethane
 Concen: 20.418 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39

Tgt Ion: 49 Resp: 28180
 Ion Ratio Lower Upper
 49 100
 129 1.3 0.0 3.4
 130 70.2 60.2 90.2





#29

Tetrahydrofuran

Concen: 97.994 ug/l

RT: 7.842 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion: 42 Resp: 66007

Ion Ratio Lower Upper

42 100

72 44.8 35.9 53.9

71 40.7 33.9 50.9



#30

Chloroform

Concen: 19.023 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN083768.D




Acq: 11 Sep 2024 03:39


Tgt Ion: 83 Resp: 66187

Ion Ratio Lower Upper

83 100

85 58.4 52.8 79.2









#31
Cyclohexane
Concen: 18.464 ug/l
RT: 8.247 min Scan# 1070
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 56 Resp: 57697
Ion Ratio Lower Upper
56 100
69 31.3 26.9 40.3
84 83.7 69.2 103.8



#32
1,1,1-Trichloroethane
Concen: 19.575 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 97 Resp: 62627
Ion Ratio Lower Upper
97 100
99 62.7 52.3 78.5
61 47.0 38.2 57.2





#33

1,2-Dichloroethane-d4

Concen: 52.208 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion: 65 Resp: 112878

Ion Ratio Lower Upper

65 100

67 50.3 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion: 114 Resp: 257368


Ion Ratio Lower Upper

114 100

63 23.6 0.0 43.4

88 17.2 0.0 32.6









#35
Dibromofluoromethane
Concen: 50.126 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:113 Resp: 82589
Ion Ratio Lower Upper
113 100
111 101.2 82.6 124.0
192 16.8 14.3 21.5



#36
1,1-Dichloropropene
Concen: 17.841 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 75 Resp: 44245
Ion Ratio Lower Upper
75 100
110 31.7 16.7 50.1
77 30.8 25.2 37.8





#37

Ethyl Acetate

Concen: 19.525 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion: 43 Resp: 43951

Ion Ratio Lower Upper

43 100

61 14.5 11.1 16.7

70 11.0 8.8 13.2



#38

Carbon Tetrachloride

Concen: 18.335 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion: 117 Resp: 51584


Ion Ratio Lower Upper

117 100

119 97.4 76.1 114.1

121 30.6 24.1 36.1





#39

Methylcyclohexane

Concen: 17.670 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion: 83 Resp: 49877

Ion Ratio Lower Upper

83 100

55 79.6 62.2 93.4

98 48.6 37.6 56.4



#40

Benzene

Concen: 18.898 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083768.D




Acq: 11 Sep 2024 03:39

Tgt Ion: 78 Resp: 141057

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.6








#41
Methacrylonitrile
Concen: 19.568 ug/l
RT: 7.771 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:	41	Resp:	26053
Ion	Ratio	Lower	Upper
41	100		
39	56.0	45.4	68.2
67	66.1	55.9	83.9
52	29.4	24.1	36.1



#42
1,2-Dichloroethane
Concen: 19.024 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:	62	Resp:	53060
Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	16.6









#43
Isopropyl Acetate
Concen: 19.683 ug/l
RT: 8.689 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion:	43	Resp:	80290
Ion	Ratio	Lower	Upper
43	100		
61	23.9	19.4	29.0
87	11.8	9.8	14.6



#44
Trichloroethene
Concen: 18.145 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:	130	Resp:	31364
Ion	Ratio	Lower	Upper
130	100		
95	96.8	0.0	200.4





#45

1,2-Dichloropropane

Concen: 18.924 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion: 63 Resp: 34140

Ion Ratio Lower Upper

63 100

65 29.7 25.9 38.9



#46

Dibromomethane

Concen: 19.371 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion: 93 Resp: 24478


Ion Ratio Lower Upper

93 100

95 80.6 65.3 97.9

174 85.4 71.0 106.4









#47
Bromodichloromethane
Concen: 18.986 ug/l
RT: 9.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion: 83 Resp: 51921
Ion Ratio Lower Upper
83 100
85 64.7 51.8 77.8
127 9.4 6.6 9.8



#48
Methyl methacrylate
Concen: 18.508 ug/l
RT: 9.683 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 41 Resp: 35820
Ion Ratio Lower Upper
41 100
69 81.5 64.3 96.5
39 66.2 48.9 73.3





#49

1,4-Dioxane

Concen: 361.588 ug/l

RT: 9.694 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion: 88 Resp: 11512

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 72.7 57.1 85.7



#50

Toluene-d8

Concen: 50.765 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN083768.D




Acq: 11 Sep 2024 03:39


Tgt Ion: 98 Resp: 289156

Ion Ratio Lower Upper

98 100

100 65.9 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 96.339 ug/l
RT: 10.447 min Scan# 1443
Delta R.T. 0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 43 Resp: 213921
Ion Ratio Lower Upper
43 100
58 38.3 31.6 47.4



#52
Toluene
Concen: 18.577 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 92 Resp: 84819
Ion Ratio Lower Upper
92 100
91 174.6 134.7 202.1









#53
t-1,3-Dichloropropene
Concen: 17.092 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 75 Resp: 47380
Ion Ratio Lower Upper
75 100
77 34.6 25.8 38.6



#54
cis-1,3-Dichloropropene
Concen: 17.969 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 75 Resp: 52323
Ion Ratio Lower Upper
75 100
77 31.8 25.0 37.6
39 57.1 42.6 64.0





#55

1,1,2-Trichloroethane

Concen: 18.849 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02

Tgt Ion: 97 Resp: 32129





Ion Ratio Lower Upper

97 100

83 80.3 67.5 101.3

85 52.8 44.2 66.2

99 53.8 50.1 75.1



#56

Ethyl methacrylate

Concen: 18.884 ug/l

RT: 10.877 min Scan# 1517

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion: 69 Resp: 49523


Ion Ratio Lower Upper

69 100

41 70.1 53.9 80.9

39 43.4 32.6 48.8









#57
1,3-Dichloropropane
Concen: 18.943 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion: 76 Resp: 55493
Ion Ratio Lower Upper
76 100
78 32.6 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 92.738 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 63 Resp: 87631
Ion Ratio Lower Upper
63 100
106 28.4 23.3 34.9









#59
2-Hexanone
Concen: 93.665 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 43 Resp: 153851
Ion Ratio Lower Upper
43 100
58 53.0 27.2 81.5



#60
Dibromochloromethane
Concen: 18.785 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 129 Resp: 35833
Ion Ratio Lower Upper
129 100
127 78.6 39.0 116.9









#61
1,2-Dibromoethane
Concen: 18.687 ug/l
RT: 11.471 min Scan# 1
Delta R.T. 0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion:107 Resp: 31767
Ion Ratio Lower Upper
107 100
109 92.5 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 47.417 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 95 Resp: 111431
Ion Ratio Lower Upper
95 100
174 69.4 0.0 139.8
176 68.4 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion:117 Resp: 227885

Ion Ratio Lower Upper

117 100

82 61.5 47.4 71.0

119 33.3 26.3 39.5



#64

Tetrachloroethene

Concen: 18.031 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Tgt Ion:164 Resp: 27325




Ion Ratio Lower Upper


164 100

166 121.5 103.7 155.5

129 100.7 79.9 119.9

131 98.8 78.4 117.6









#65
Chlorobenzene
Concen: 17.456 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:112 Resp: 88414
Ion Ratio Lower Upper
112 100
114 31.9 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 17.986 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:131 Resp: 31580
Ion Ratio Lower Upper
131 100
133 98.8 47.5 142.6
119 64.8 31.6 94.8









#67
Ethyl Benzene
Concen: 17.779 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion: 91 Resp: 159960
Ion Ratio Lower Upper
91 100
106 28.4 24.3 36.5



#68
m/p-Xylenes
Concen: 35.284 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:106 Resp: 117861
Ion Ratio Lower Upper
106 100
91 212.6 169.2 253.8





#69

o-Xylene

Concen: 18.365 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :





VN0910WBS02

Tgt Ion:106 Resp: 59388

Ion Ratio Lower Upper

106 100

91 220.1 112.9 338.6



#70

Styrene

Concen: 18.154 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion:104 Resp: 96345


Ion Ratio Lower Upper

104 100

78 54.7 43.3 64.9

103 53.7 43.4 65.2





#71

Bromoform

Concen: 18.369 ug/l

RT: 12.576 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion:173 Resp: 21734

Ion Ratio Lower Upper

173 100

175 49.3 24.1 72.2

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion:152 Resp: 108516


Ion Ratio Lower Upper

152 100

115 64.5 32.2 96.6

150 161.4 0.0 348.6









#73
Isopropylbenzene
Concen: 17.792 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:105 Resp: 149372
Ion Ratio Lower Upper
105 100
120 26.1 12.6 37.8



#74
N-ethyl acetate
Concen: 18.128 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 43 Resp: 64934
Ion Ratio Lower Upper
43 100
70 41.5 34.5 51.7
55 25.4 18.6 28.0
61 23.8 19.2 28.8









#75
 1,1,2,2-Tetrachloroethane
 Concen: 18.333 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02


Tgt Ion: 83 Resp: 45224
 Ion Ratio Lower Upper
 83 100
 131 9.6 5.2 15.6
 85 63.3 31.9 95.9



#76
 1,2,3-Trichloropropane
 Concen: 25.813 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39

Tgt Ion: 75 Resp: 56520
 Ion Ratio Lower Upper
 75 100
 77 142.6 95.2 285.6









#77
Bromobenzene
Concen: 16.868 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion:156 Resp: 34088
Ion Ratio Lower Upper
156 100
77 236.4 114.3 343.1
158 99.0 48.9 146.7



#78
n-propylbenzene
Concen: 17.519 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 91 Resp: 171132
Ion Ratio Lower Upper
91 100
120 21.7 11.0 33.0









#79
2-Chlorotoluene
Concen: 17.820 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion: 91 Resp: 111010
Ion Ratio Lower Upper
91 100
126 29.9 15.4 46.2



#80
1,3,5-Trimethylbenzene
Concen: 18.102 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:105 Resp: 128127
Ion Ratio Lower Upper
105 100
120 45.6 23.4 70.3









#81
trans-1,4-Dichloro-2-butene
Concen: 16.120 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02

Tgt Ion: 75 Resp: 13477
Ion Ratio Lower Upper
75 100
53 107.2 81.2 121.8
89 44.3 35.7 53.5



#82
4-Chlorotoluene
Concen: 17.086 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion: 91 Resp: 106816
Ion Ratio Lower Upper
91 100
126 31.1 15.4 46.2





#83

tert-Butylbenzene

Concen: 17.976 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion:119 Resp: 111318

Ion Ratio Lower Upper

119 100

91 73.0 34.5 103.6

134 26.6 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 17.831 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN083768.D




Acq: 11 Sep 2024 03:39


Tgt Ion:105 Resp: 126783

Ion Ratio Lower Upper

105 100

120 43.2 21.8 65.4









#85
 sec-Butylbenzene
 Concen: 17.766 ug/l
 RT: 13.612 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0910WBS02


Tgt Ion:105 Resp: 145907
 Ion Ratio Lower Upper
 105 100
 134 18.7 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 17.224 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN083768.D
 Acq: 11 Sep 2024 03:39

Tgt Ion:119 Resp: 118583
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.8 38.3
 91 26.2 12.9 38.7









#87
1,3-Dichlorobenzene
Concen: 16.521 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:146 Resp: 63010
Ion Ratio Lower Upper
146 100
111 44.1 21.7 65.1
148 64.6 31.7 95.1



#88
1,4-Dichlorobenzene
Concen: 16.258 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:146 Resp: 62787
Ion Ratio Lower Upper
146 100
111 44.1 21.3 64.0
148 63.6 32.0 96.2









#89
n-Butylbenzene
Concen: 15.552 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:	91	Resp:	96128
Ion	Ratio	Lower	Upper
91	100		
92	53.5	26.7	80.1
134	24.8	12.2	36.4



#90
Hexachloroethane
Concen: 17.880 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:	117	Resp:	24745
Ion	Ratio	Lower	Upper
117	100		
201	63.8	32.8	98.3





#91

1,2-Dichlorobenzene

Concen: 17.329 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion:146 Resp: 63786

Ion Ratio Lower Upper

146 100

111 45.4 22.8 68.3

148 63.7 32.1 96.5



#92

1,2-Dibromo-3-Chloropropane

Concen: 17.853 ug/l

RT: 14.718 min Scan# 2170

Delta R.T. 0.000 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39




Tgt Ion: 75 Resp: 8779


Ion Ratio Lower Upper

75 100

155 73.9 35.9 107.7

157 88.4 47.5 142.5









#93
1,2,4-Trichlorobenzene
Concen: 15.171 ug/l
RT: 15.388 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39




Instrument :
MSVOA_N
ClientSampleId :
VN0910WBS02


Tgt Ion:	180	Resp:	29965
Ion Ratio	Lower	Upper	
180	100		
182	95.5	48.1	144.3
145	35.4	16.9	50.7



#94
Hexachlorobutadiene
Concen: 15.552 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN083768.D
Acq: 11 Sep 2024 03:39

Tgt Ion:	225	Resp:	13045
Ion Ratio	Lower	Upper	
225	100		
223	62.5	31.9	95.5
227	59.8	32.0	96.2





#95

Naphthalene

Concen: 16.071 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Instrument :

MSVOA_N

ClientSampleId :

VN0910WBS02





Tgt Ion:128 Resp: 96573

Ion Ratio Lower Upper

128 100

127 12.8 10.2 15.2

129 10.6 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 15.627 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.006 min

Lab File: VN083768.D

Acq: 11 Sep 2024 03:39

Tgt Ion:180 Resp: 29500

Ion Ratio Lower Upper

180 100

182 95.9 47.1 141.4

145 36.0 17.8 53.5

