

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083773.D
 Acq On : 11 Sep 2024 05:38
 Operator : JC\MD
 Sample : PB163204ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163204ZHE#03

Quant Time: Sep 11 06:14:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 03:05:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.218	168	98266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	215943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118854	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	100759	68.002	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 136.000%#			
35) Dibromofluoromethane	8.159	113	79016	57.157	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 114.320%			
50) Toluene-d8	10.565	98	295916	61.918	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 123.840%#			
62) 4-Bromofluorobenzene	12.847	95	134082	68.001	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 136.000%#			
Target Compounds						Qvalue
3) Chloromethane	2.359	50	996	Below Cal		93
6) Chloroethane	3.159	64	7742	8.921 ug/l #		45
13) Acrolein	4.177	56	54	0.276 ug/l #		3
16) Acetone	4.436	43	29455	49.589 ug/l		96
17) Carbon Disulfide	4.706	76	740	0.249 ug/l #		75
18) Methyl Acetate	5.018	43	8779	5.440 ug/l #		84
20) Methylene Chloride	5.259	84	6257	5.148 ug/l #		76
25) 2-Butanone	7.488	43	3608	4.622 ug/l #		75
29) Tetrahydrofuran	7.853	42	198	0.429 ug/l #		32
31) Cyclohexane	8.218	56	2481	1.159 ug/l #		35
36) 1,1-Dichloropropene	8.218	75	9179	4.411 ug/l #		48
37) Ethyl Acetate	7.565	43	2969	1.572 ug/l #		84
38) Carbon Tetrachloride	8.224	117	11373	4.818 ug/l #		18
43) Isopropyl Acetate	8.688	43	156067	45.599 ug/l #		89
45) 1,2-Dichloropropane	9.747	63	1239	0.819 ug/l #		42
48) Methyl methacrylate	9.741	41	103411	63.683 ug/l #		46
51) 4-Methyl-2-Pentanone	10.441	43	5095	2.735 ug/l		95
53) t-1,3-Dichloropropene	10.724	75	510	0.219 ug/l #		42
54) cis-1,3-Dichloropropene	10.353	75	41213	16.869 ug/l #		65
56) Ethyl methacrylate	10.988	69	1139	0.518 ug/l #		1
57) 1,3-Dichloropropane	11.200	76	5317	2.163 ug/l #		42
59) 2-Hexanone	11.171	43	105094	76.255 ug/l #		24
74) N-amyl acetate	12.476	43	1322	0.337 ug/l #		43
76) 1,2,3-Trichloropropane	12.847	75	75838	31.622 ug/l #		1
81) trans-1,4-Dichloro-2-b...	12.847	75	75838	82.821 ug/l #		9

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083773.D
Acq On : 11 Sep 2024 05:38
Operator : JC\MD
Sample : PB163204ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Quant Time: Sep 11 06:14:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 03:05:46 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.006 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03







Tgt Ion:168 Resp: 98266
Ion Ratio Lower Upper
168 100
99 69.9 54.2 81.2



#3
Chloromethane
Concen: Below Cal
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 50 Resp: 996
Ion Ratio Lower Upper
50 100
52 38.1 27.2 40.8









#6
Chloroethane
Concen: 8.921 ug/l
RT: 3.159 min Scan# 205
Delta R.T. 0.076 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38




Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Tgt Ion: 64 Resp: 7742
Ion Ratio Lower Upper
64 100
66 0.0 23.9 35.9#



#13
Acrolein
Concen: 0.276 ug/l
RT: 4.177 min Scan# 378
Delta R.T. 0.006 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 56 Resp: 54
Ion Ratio Lower Upper
56 100
55 153.7 58.2 87.2#





#16

Acetone

Concen: 49.589 ug/l

RT: 4.436 min Scan# 41

Delta R.T. 0.006 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38

Instrument :

MSVOA_N

ClientSampleId :





PB163204ZHE#03

Tgt Ion: 43 Resp: 29455

Ion Ratio Lower Upper

43 100

58 31.8 23.7 35.5



#17

Carbon Disulfide

Concen: 0.249 ug/l

RT: 4.706 min Scan# 468

Delta R.T. 0.006 min

Lab File: VN083773.D




Acq: 11 Sep 2024 05:38

Tgt Ion: 76 Resp: 740

Ion Ratio Lower Upper

76 100

78 0.0 7.1 10.7#









#18
Methyl Acetate
Concen: 5.440 ug/l
RT: 5.018 min Scan# 51
Delta R.T. -0.006 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38




Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Tgt Ion: 43 Resp: 8779
Ion Ratio Lower Upper
43 100
74 15.8 19.0 28.4#



#20
Methylene Chloride
Concen: 5.148 ug/l
RT: 5.259 min Scan# 562
Delta R.T. -0.012 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 84 Resp: 6257
Ion Ratio Lower Upper
84 100
49 139.5 93.4 140.0
51 59.1 29.2 43.8#
86 74.9 46.9 70.3#





#25

2-Butanone

Concen: 4.622 ug/l

RT: 7.488 min Scan# 940

Delta R.T. 0.006 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38

Instrument :

MSVOA_N

ClientSampleId :





PB163204ZHE#03

Tgt Ion: 43 Resp: 3608

Ion Ratio Lower Upper

43 100

72 13.3 20.8 31.2#



#29

Tetrahydrofuran

Concen: 0.429 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.018 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38




Tgt Ion: 42 Resp: 198


Ion Ratio Lower Upper

42 100

72 0.0 35.9 53.9#


71 0.0 33.9 50.9#








#31
Cyclohexane
Concen: 1.159 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.035 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03







Tgt Ion: 56 Resp: 2481
Ion Ratio Lower Upper
56 100
69 155.5 26.9 40.3#
84 92.8 69.2 103.8



#33
1,2-Dichloroethane-d4
Concen: 68.002 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 65 Resp: 100759
Ion Ratio Lower Upper
65 100
67 50.2 0.0 100.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38

Instrument :

MSVOA_N

ClientSampleId :

PB163204ZHE#03





Tgt Ion:114 Resp: 215943

Ion Ratio Lower Upper

114 100

63 22.3 0.0 43.4

88 16.8 0.0 32.6



#35

Dibromofluoromethane

Concen: 57.157 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. -0.006 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38




Tgt Ion:113 Resp: 79016


Ion Ratio Lower Upper

113 100

111 102.7 82.6 124.0


192 17.5 14.3 21.5








#36
1,1-Dichloropropene
Concen: 4.411 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.153 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03







Tgt Ion: 75 Resp: 9179
Ion Ratio Lower Upper
75 100
110 6.2 16.7 50.1#
77 0.0 25.2 37.8#



#37
Ethyl Acetate
Concen: 1.572 ug/l
RT: 7.565 min Scan# 954
Delta R.T. 0.006 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38


Tgt Ion: 43 Resp: 2969
Ion Ratio Lower Upper
43 100
61 9.3 11.1 16.7#
70 2.4 8.8 13.2#








#38
Carbon Tetrachloride
Concen: 4.818 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.135 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03







Tgt Ion:117 Resp: 11373
Ion Ratio Lower Upper
117 100
119 6.8 76.1 114.1#
121 0.0 24.1 36.1#



#43
Isopropyl Acetate
Concen: 45.599 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 43 Resp: 156067
Ion Ratio Lower Upper
43 100
61 17.1 19.4 29.0#
87 10.0 9.8 14.6





#45

1,2-Dichloropropane

Concen: 0.819 ug/l

RT: 9.747 min Scan# 11

Delta R.T. 0.129 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38

Instrument :

MSVOA_N

ClientSampleId :





PB163204ZHE#03

Tgt Ion: 63 Resp: 1239

Ion Ratio Lower Upper

63 100

65 0.0 25.9 38.9#



#48

Methyl methacrylate

Concen: 63.683 ug/l

RT: 9.741 min Scan# 1324

Delta R.T. 0.065 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38




Tgt Ion: 41 Resp: 103411


Ion Ratio Lower Upper

41 100

69 0.0 64.3 96.5#

39 57.1 48.9 73.3





#50

Toluene-d8

Concen: 61.918 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083773.D

Acq: 11 Sep 2024 05:38

Instrument :

MSVOA_N

ClientSampleId :





PB163204ZHE#03

Tgt Ion: 98 Resp: 295916

Ion Ratio Lower Upper

98 100

100 64.3 52.0 78.0



#51

4-Methyl-2-Pentanone

Concen: 2.735 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. -0.000 min

Lab File: VN083773.D




Acq: 11 Sep 2024 05:38


Tgt Ion: 43 Resp: 5095

Ion Ratio Lower Upper

43 100

58 36.2 31.6 47.4









#53
 t-1,3-Dichloropropene
 Concen: 0.219 ug/l
 RT: 10.724 min Scan# 1491
 Delta R.T. -0.112 min
 Lab File: VN083773.D
 Acq: 11 Sep 2024 05:38




Instrument :
 MSVOA_N
 ClientSampleId :
 PB163204ZHE#03


Tgt Ion: 75 Resp: 510
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.8 38.6#



#54
 cis-1,3-Dichloropropene
 Concen: 16.869 ug/l
 RT: 10.353 min Scan# 1428
 Delta R.T. 0.041 min
 Lab File: VN083773.D
 Acq: 11 Sep 2024 05:38

Tgt Ion: 75 Resp: 41213
 Ion Ratio Lower Upper
 75 100
 77 0.8 25.0 37.6#
 39 36.5 42.6 64.0#









#56
Ethyl methacrylate
Concen: 0.518 ug/l
RT: 10.988 min Scan# 11
Delta R.T. 0.118 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38




Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Tgt Ion: 69 Resp: 1139
Ion Ratio Lower Upper
69 100
41 6076.9 53.9 80.9#
39 2650.0 32.6 48.8#



#57
1,3-Dichloropropane
Concen: 2.163 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.041 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 76 Resp: 5317
Ion Ratio Lower Upper
76 100
78 0.0 25.8 38.6#









#59
2-Hexanone
Concen: 76.255 ug/l
RT: 11.171 min Scan# 1567
Delta R.T. -0.024 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38




Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Tgt Ion: 43 Resp: 105094
Ion Ratio Lower Upper
43 100
58 0.0 27.2 81.5#



#62
4-Bromofluorobenzene
Concen: 68.001 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 95 Resp: 134082
Ion Ratio Lower Upper
95 100
174 70.5 0.0 139.8
176 67.8 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38




Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Tgt Ion:117 Resp: 250985
Ion Ratio Lower Upper
117 100
82 61.0 47.4 71.0
119 31.6 26.3 39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion:152 Resp: 118854
Ion Ratio Lower Upper
152 100
115 64.6 32.2 96.6
150 162.3 0.0 348.6









#74
N-amy1 acetate
Concen: 0.337 ug/l
RT: 12.476 min Scan# 11
Delta R.T. -0.018 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38




Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#03


Tgt Ion: 43	Resp:	1322	
Ion	Ratio	Lower	Upper
43	100		
70	0.0	34.5	51.7#
55	0.0	18.6	28.0#
61	0.0	19.2	28.8#



#76
1,2,3-Trichloropropane
Concen: 31.622 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.147 min
Lab File: VN083773.D
Acq: 11 Sep 2024 05:38

Tgt Ion: 75	Resp: 75838
Ion Ratio	Lower Upper
75 100	
77 1.2	95.2 285.6#





#81
 trans-1,4-Dichloro-2-butene
 Concen: 82.821 ug/l
 RT: 12.847 min Scan# 11
 Delta R.T. 0.112 min
 Lab File: VN083773.D
 Acq: 11 Sep 2024 05:38

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163204ZHE#03

Tgt Ion: 75 Resp: 75838
 Ion Ratio Lower Upper
 75 100
 53 0.0 81.2 121.8#
 89 0.0 35.7 53.5#

