

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083756.D
 Acq On : 10 Sep 2024 16:55
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 11 03:16:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 02:47:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	107993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	221590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108324	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	88247	54.193	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.380%
35) Dibromofluoromethane	8.165	113	69503	48.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.980%
50) Toluene-d8	10.565	98	249806	50.938	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.880%
62) 4-Bromofluorobenzene	12.847	95	109816	54.275	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.540%
Target Compounds						
					Qvalue	
53) t-1,3-Dichloropropene	10.829	75	3052	1.279	ug/l #	83
70) Styrene	12.406	104	6837	1.238	ug/l	95
87) 1,3-Dichlorobenzene	13.735	146	9635	2.531	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	542	1.104	ug/l	88
95) Naphthalene	15.641	128	17076	2.847	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083756.D
Acq On : 10 Sep 2024 16:55
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 11 03:16:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 02:47:22 2024
Response via : Initial Calibration

