

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083757.D
 Acq On : 10 Sep 2024 17:20
 Operator : JC\MD
 Sample : P3841-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OW-07B-65.5-090424DL

Quant Time: Sep 11 03:16:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 02:47:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	116124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	242469	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	257703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	94265	53.835	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.680%	
35) Dibromofluoromethane	8.165	113	74695	48.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.240%	
50) Toluene-d8	10.565	98	261217	48.678	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.360%	
62) 4-Bromofluorobenzene	12.847	95	120499	54.427	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.860%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.324	96	27114	21.588	ug/l	92
27) cis-1,2-Dichloroethene	7.483	96	22265	13.934	ug/l	88
44) Trichloroethene	9.347	130	25836	15.865	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083757.D
Acq On : 10 Sep 2024 17:20
Operator : JC\MD
Sample : P3841-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OW-07B-65.5-090424DL

Quant Time: Sep 11 03:16:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 02:47:22 2024
Response via : Initial Calibration

