

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083759.D
 Acq On : 10 Sep 2024 18:08
 Operator : JC\MD
 Sample : P3868-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OWBR-06-219-090524

Quant Time: Sep 11 03:17:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 02:47:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	112306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	231588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	91932	54.288	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.580%	
35) Dibromofluoromethane	8.165	113	73153	49.341	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.680%	
50) Toluene-d8	10.565	98	256735	50.091	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.180%	
62) 4-Bromofluorobenzene	12.847	95	117704	55.662	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 111.320%	
Target Compounds						
52) Toluene	10.629	92	16740	4.075	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083759.D
Acq On : 10 Sep 2024 18:08
Operator : JC\MD
Sample : P3868-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OWBR-06-219-090524

Quant Time: Sep 11 03:17:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 02:47:22 2024
Response via : Initial Calibration

