

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083782.D
 Acq On : 11 Sep 2024 09:14
 Operator : JC\MD
 Sample : PB163204ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163204ZHE#12

Quant Time: Sep 11 09:41:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 03:05:46 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	93213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	207876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	96679	68.785	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 137.580%#	
35) Dibromofluoromethane	8.165	113	77463	58.208	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.420%	
50) Toluene-d8	10.565	98	288781	62.770	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 125.540%#	
62) 4-Bromofluorobenzene	12.847	95	134453	70.835	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 141.680%#	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	34923	61.982	ug/l	91
18) Methyl Acetate	5.024	43	9747	6.696	ug/l	93
20) Methylene Chloride	5.265	84	8642	7.495	ug/l	88
25) 2-Butanone	7.488	43	4127	5.573	ug/l	99
43) Isopropyl Acetate	8.688	43	262394	79.640	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083782.D
Acq On : 11 Sep 2024 09:14
Operator : JC\MD
Sample : PB163204ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163204ZHE#12

Quant Time: Sep 11 09:41:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 03:05:46 2024
Response via : Initial Calibration

