

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
 Data File : VN083787.D
 Acq On : 11 Sep 2024 11:14
 Operator : JC\MD
 Sample : P3906-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-701-COMP-03

Quant Time: Sep 11 11:33:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 03:05:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	111837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	246622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	280899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	99368	58.925	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.860%
35) Dibromofluoromethane	8.165	113	76733	48.601	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.200%
50) Toluene-d8	10.565	98	283820	51.999	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.000%
62) 4-Bromofluorobenzene	12.847	95	132622	58.893	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	117.780%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	39805	58.882	ug/l	99
18) Methyl Acetate	5.030	43	9140	4.812	ug/l	99
20) Methylene Chloride	5.271	84	6519	4.712	ug/l	89
43) Isopropyl Acetate	8.688	43	100510	25.713	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091024\
Data File : VN083787.D
Acq On : 11 Sep 2024 11:14
Operator : JC\MD
Sample : P3906-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-701-COMP-03

Quant Time: Sep 11 11:33:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 03:05:46 2024
Response via : Initial Calibration

