

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091025\
 Data File : VN087785.D
 Acq On : 10 Sep 2025 16:50
 Operator : JC\MD
 Sample : Q3045-05DL 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DNAPL-OILDL

Quant Time: Sep 11 02:07:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.206	168	188752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	444979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	418342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	209155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	194757	50.360	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	100.720%	
35) Dibromofluoromethane	8.147	113	156894	48.835	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	97.660%	
50) Toluene-d8	10.547	98	556657	46.293	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	92.580%	
62) 4-Bromofluorobenzene	12.829	95	202783	47.420	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	=	94.840%	
Target Compounds						
					Qvalue	
5) Bromomethane	2.989	94	619	0.376	ug/l #	77
10) Methyl Iodide	4.594	142	59	6.836	ug/l #	38
13) Acrolein	4.183	56	198	0.240	ug/l #	1
16) Acetone	4.424	43	6383	3.032	ug/l	91
20) Methylene Chloride	5.265	84	175	Below Cal	#	8
25) 2-Butanone	7.482	43	4771	1.723	ug/l #	86
31) Cyclohexane	8.206	56	5545	1.140	ug/l #	50
36) 1,1-Dichloropropene	8.206	75	18301	3.713	ug/l #	51
38) Carbon Tetrachloride	8.212	117	21821	4.484	ug/l #	14
40) Benzene	8.588	78	2096857	138.707	ug/l	100
51) 4-Methyl-2-Pentanone	10.341	43	1479	0.233	ug/l #	38
52) Toluene	10.612	92	1020684	108.910	ug/l	99
53) t-1,3-Dichloropropene	10.606	75	13038	2.167	ug/l #	1
55) 1,1,2-Trichloroethane	10.865	97	5999	1.655	ug/l #	18
67) Ethyl Benzene	11.941	91	429316	23.824	ug/l	99
68) m/p-Xylenes	12.047	106	194251	29.393	ug/l	99
69) o-Xylene	12.376	106	106200	16.398	ug/l	97
70) Styrene	12.388	104	166238	15.321	ug/l	96
71) Bromoform	12.823	173	1023	0.386	ug/l #	1
73) Isopropylbenzene	12.676	105	10606	0.628	ug/l	88
76) 1,2,3-Trichloropropane	12.829	75	114792	23.404	ug/l #	1
78) n-propylbenzene	13.017	91	7961	0.382	ug/l	99
79) 2-Chlorotoluene	13.070	91	5828	0.465	ug/l #	44
80) 1,3,5-Trimethylbenzene	13.147	105	21022	1.466	ug/l	84
81) trans-1,4-Dichloro-2-b...	12.829	75	114792	61.960	ug/l #	1
83) tert-Butylbenzene	13.459	119	10238	0.857	ug/l #	65
84) 1,2,4-Trimethylbenzene	13.459	105	77284	5.385	ug/l	100
85) sec-Butylbenzene	13.459	105	77284	4.359	ug/l #	58
86) p-Isopropyltoluene	13.706	119	2933	0.200	ug/l #	84
90) Hexachloroethane	14.358	117	4077	1.462	ug/l #	22
95) Naphthalene	15.611	128	2841260	179.426	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091025\
Data File : VN087785.D
Acq On : 10 Sep 2025 16:50
Operator : JC\MD
Sample : Q3045-05DL 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL

Quant Time: Sep 11 02:07:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion:168 Resp: 188752
Ion Ratio Lower Upper
168 100
99 68.4 57.2 85.8



#5
Bromomethane
Concen: 0.376 ug/l
RT: 2.989 min Scan# 176
Delta R.T. 0.017 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50


Tgt Ion: 94 Resp: 619
Ion Ratio Lower Upper
94 100
96 72.4 75.4 113.2#








#10
Methyl Iodide
Concen: 6.836 ug/l
RT: 4.594 min Scan# 446
Delta R.T. 0.017 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL







Tgt Ion:142 Resp: 59
Ion Ratio Lower Upper
142 100
127 0.0 38.0 57.0#
141 0.0 12.0 18.0#



#13
Acrolein
Concen: 0.240 ug/l
RT: 4.183 min Scan# 379
Delta R.T. 0.012 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 56 Resp: 198
Ion Ratio Lower Upper
56 100
55 269.2 57.3 85.9#









#16
Acetone
Concen: 3.032 ug/l
RT: 4.424 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion: 43 Resp: 6383
Ion Ratio Lower Upper
43 100
58 23.4 22.6 33.8



#20
Methylene Chloride
Concen: Below Cal
RT: 5.265 min Scan# 563
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 84 Resp: 175
Ion Ratio Lower Upper
84 100
49 0.0 103.4 155.2#
51 74.1 33.8 50.8#
86 0.0 49.6 74.4#









#25
2-Butanone
Concen: 1.723 ug/l
RT: 7.482 min Scan# 940
Delta R.T. 0.017 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion: 43 Resp: 4771
Ion Ratio Lower Upper
43 100
72 17.9 19.9 29.9#



#31
Cyclohexane
Concen: 1.140 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. -0.030 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 56 Resp: 5545
Ion Ratio Lower Upper
56 100
69 120.5 23.7 35.5#
84 86.2 64.7 97.1





#33

1,2-Dichloroethane-d4

Concen: 50.360 ug/l

RT: 8.565 min Scan# 1123

Delta R.T. 0.006 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50

Instrument :

MSVOA_N

ClientSampleId :





DNAPL-OILDL

Tgt Ion: 65 Resp: 194757

Ion Ratio Lower Upper

65 100

67 51.9 0.0 100.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50




Tgt Ion: 114 Resp: 444979


Ion Ratio Lower Upper

114 100

63 22.0 0.0 45.2

88 15.9 0.0 33.4





#35

Dibromofluoromethane

Concen: 48.835 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50

Instrument :

MSVOA_N

ClientSampleId :

DNAPL-OILDL





Tgt Ion:113 Resp: 156894

Ion Ratio Lower Upper

113 100

111 99.5 82.8 124.2

192 16.4 12.8 19.2



#36

1,1-Dichloropropene

Concen: 3.713 ug/l

RT: 8.206 min Scan# 1063

Delta R.T. -0.147 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50




Tgt Ion: 75 Resp: 18301


Ion Ratio Lower Upper

75 100

110 8.5 15.9 47.6#


77 0.0 24.4 36.6#








#38
Carbon Tetrachloride
Concen: 4.484 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.130 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL







Tgt Ion:117 Resp: 21821
Ion Ratio Lower Upper
117 100
119 4.6 79.0 118.6#
121 0.0 25.0 37.6#



#40
Benzene
Concen: 138.707 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 78 Resp: 2096857
Ion Ratio Lower Upper
78 100
77 24.6 19.7 29.5









#50
Toluene-d8
Concen: 46.293 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion: 98 Resp: 556657
Ion Ratio Lower Upper
98 100
100 63.7 52.8 79.2



#51
4-Methyl-2-Pentanone
Concen: 0.233 ug/l
RT: 10.341 min Scan# 1426
Delta R.T. -0.083 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 43 Resp: 1479
Ion Ratio Lower Upper
43 100
58 0.0 29.6 44.4#









#52
Toluene
Concen: 108.910 ug/l
RT: 10.612 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion: 92 Resp: 1020684
Ion Ratio Lower Upper
92 100
91 174.4 140.4 210.6



#53
t-1,3-Dichloropropene
Concen: 2.167 ug/l
RT: 10.606 min Scan# 1471
Delta R.T. -0.212 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 75 Resp: 13038
Ion Ratio Lower Upper
75 100
77 169.0 25.2 37.8#









#55
 1,1,2-Trichloroethane
 Concen: 1.655 ug/l
 RT: 10.865 min Scan# 11
 Delta R.T. -0.130 min
 Lab File: VN087785.D
 Acq: 10 Sep 2025 16:50




Instrument :
 MSVOA_N
 ClientSampleId :
 DNAPL-OILDL


Tgt Ion: 97	Resp: 5999
Ion Ratio	Lower Upper
97 100	
83 0.0	68.2 102.4#
85 0.0	45.5 68.3#
99 9.9	52.2 78.4#



#62
 4-Bromofluorobenzene
 Concen: 47.420 ug/l
 RT: 12.829 min Scan# 1849
 Delta R.T. 0.006 min
 Lab File: VN087785.D
 Acq: 10 Sep 2025 16:50

Tgt Ion: 95	Resp: 202783
Ion Ratio	Lower Upper
95 100	
174 72.5	0.0 138.4
176 70.2	0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion:117 Resp: 418342
Ion Ratio Lower Upper
117 100
82 54.3 47.4 71.2
119 29.9 24.3 36.5



#67
Ethyl Benzene
Concen: 23.824 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion: 91 Resp: 429316
Ion Ratio Lower Upper
91 100
106 29.7 24.1 36.1









#68
m/p-Xylenes
Concen: 29.393 ug/l
RT: 12.047 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion:106 Resp: 194251
Ion Ratio Lower Upper
106 100
91 210.2 169.3 253.9



#69
o-Xylene
Concen: 16.398 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion:106 Resp: 106200
Ion Ratio Lower Upper
106 100
91 228.2 111.4 334.2





#70

Styrene

Concen: 15.321 ug/l

RT: 12.388 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50

Instrument :

MSVOA_N

ClientSampleId :

DNAPL-OILDL





Tgt Ion:104 Resp: 166238

Ion Ratio Lower Upper

104 100

78 51.2 42.2 63.2

103 58.0 43.4 65.2



#71

Bromoform

Concen: 0.386 ug/l

RT: 12.823 min Scan# 1848

Delta R.T. 0.264 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50




Tgt Ion:173 Resp: 1023


Ion Ratio Lower Upper

173 100

175 901.2 24.9 74.6#

254 0.0 0.1 0.1#









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2008
Delta R.T. 0.006 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion:152 Resp: 209155
Ion Ratio Lower Upper
152 100
115 63.7 32.3 96.8
150 158.9 0.0 351.0



#73
Isopropylbenzene
Concen: 0.628 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.006 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion:105 Resp: 10606
Ion Ratio Lower Upper
105 100
120 31.4 12.8 38.3





#76

1,2,3-Trichloropropane

Concen: 23.404 ug/l

RT: 12.829 min Scan# 11

Delta R.T. -0.141 min

Lab File: VN087785.D

Acq: 10 Sep 2025 16:50

Instrument :

MSVOA_N

ClientSampleId :





DNAPL-OILDL

Tgt Ion: 75 Resp: 114792

Ion Ratio Lower Upper

75 100

77 2.3 89.1 267.4#



#78

n-propylbenzene

Concen: 0.382 ug/l

RT: 13.017 min Scan# 1881

Delta R.T. 0.006 min

Lab File: VN087785.D




Acq: 10 Sep 2025 16:50


Tgt Ion: 91 Resp: 7961

Ion Ratio Lower Upper

91 100

120 22.0 10.7 32.1









#79
 2-Chlorotoluene
 Concen: 0.465 ug/l
 RT: 13.070 min Scan# 1890
 Delta R.T. -0.030 min
 Lab File: VN087785.D
 Acq: 10 Sep 2025 16:50




Instrument :
 MSVOA_N
 ClientSampleId :
 DNAPL-OILDL


Tgt Ion: 91 Resp: 5828
 Ion Ratio Lower Upper
 91 100
 126 1.1 16.1 48.3#



#80
 1,3,5-Trimethylbenzene
 Concen: 1.466 ug/l
 RT: 13.147 min Scan# 1903
 Delta R.T. -0.006 min
 Lab File: VN087785.D
 Acq: 10 Sep 2025 16:50

Tgt Ion: 105 Resp: 21022
 Ion Ratio Lower Upper
 105 100
 120 57.1 23.2 69.6









#81
 trans-1,4-Dichloro-2-butene
 Concen: 61.960 ug/l
 RT: 12.829 min Scan# 11
 Delta R.T. 0.112 min
 Lab File: VN087785.D
 Acq: 10 Sep 2025 16:50




Instrument :
 MSVOA_N
 ClientSampleId :
 DNAPL-OILDL


Tgt Ion: 75 Resp: 114792
 Ion Ratio Lower Upper
 75 100
 53 0.0 101.8 152.6#
 89 0.0 35.6 53.4#



#83
 tert-Butylbenzene
 Concen: 0.857 ug/l
 RT: 13.459 min Scan# 1956
 Delta R.T. 0.041 min
 Lab File: VN087785.D
 Acq: 10 Sep 2025 16:50

Tgt Ion: 119 Resp: 10238
 Ion Ratio Lower Upper
 119 100
 91 94.5 35.0 105.0
 134 0.0 12.6 37.8#









#84
1,2,4-Trimethylbenzene
Concen: 5.385 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion:105 Resp: 77284
Ion Ratio Lower Upper
105 100
120 43.8 22.0 66.0



#85
sec-Butylbenzene
Concen: 4.359 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.135 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion:105 Resp: 77284
Ion Ratio Lower Upper
105 100
134 0.0 9.6 28.6#









#86
p-Isopropyltoluene
Concen: 0.200 ug/l
RT: 13.706 min Scan# 1998
Delta R.T. -0.000 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50




Instrument :
MSVOA_N
ClientSampleId :
DNAPL-OILDL


Tgt Ion:119 Resp: 2933
Ion Ratio Lower Upper
119 100
134 27.1 12.7 38.0
91 39.6 12.5 37.5#



#90
Hexachloroethane
Concen: 1.462 ug/l
RT: 14.358 min Scan# 2109
Delta R.T. 0.047 min
Lab File: VN087785.D
Acq: 10 Sep 2025 16:50

Tgt Ion:117 Resp: 4077
Ion Ratio Lower Upper
117 100
201 0.0 29.1 87.4#





#95

Naphthalene

Concen: 179.426 ug/l

RT: 15.611 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN087785.D


Acq: 10 Sep 2025 16:50

Instrument :

MSVOA_N

ClientSampleId :

DNAPL-OILDL



Tgt Ion:128 Resp: 2841260

Ion Ratio Lower Upper

128 100

127 12.9 10.6 15.8

129 10.8 8.6 12.8

