

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091118\
 Data File : VN051114.D
 Acq On : 11 Sep 2018 9:52
 Operator : MD\SY
 Sample : J4849-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-TB01

Quant Time: Sep 12 10:03:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	666851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1003981	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	864618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	283754	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	382946	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	7.59	113	380770	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	10.09	98	1408235	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
62) 4-Bromofluorobenzene	12.40	95	382444	40.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.78%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091118\
Data File : VN051114.D
Acq On : 11 Sep 2018 9:52
Operator : MD\SY
Sample : J4849-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-TB01

Quant Time: Sep 12 10:03:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

