

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057870.D
 Acq On : 10 Sep 2019 18:08
 Operator : JC/SP
 Sample : K4623-25
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-C2-(4)

Quant Time: Sep 11 03:55:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	290677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	489076	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	450593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	229594	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	328619	55.82	ug/l	0.00
Spiked Amount			Recovery	=	111.64%	
35) Dibromofluoromethane	7.58	113	219527	48.63	ug/l	0.00
Spiked Amount			Recovery	=	97.26%	
50) Toluene-d8	10.09	98	905286	51.65	ug/l	0.00
Spiked Amount			Recovery	=	103.30%	
62) 4-Bromofluorobenzene	12.40	95	291883	43.70	ug/l	0.00
Spiked Amount			Recovery	=	87.40%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.55	94	111	Below Cal	#	2
11) Tert butyl alcohol	4.72	59	491	0.689	ug/l #	72
16) Acetone	3.80	43	10523	5.452	ug/l	100
17) Carbon Disulfide	4.03	76	435	0.249	ug/l #	75
18) Methyl Acetate	4.32	43	7219	1.403	ug/l #	67
20) Methylene Chloride	4.53	84	6546	0.565	ug/l	86
29) Tetrahydrofuran	7.20	42	348	0.202	ug/l #	41
31) Cyclohexane	7.65	56	7757	Below Cal	#	43
36) 1,1-Dichloropropene	7.65	75	25490	4.494	ug/l #	50
38) Carbon Tetrachloride	7.65	117	28882	5.577	ug/l #	16
43) Isopropyl Acetate	8.16	43	91382	9.273	ug/l #	89
45) 1,2-Dichloropropane	9.26	63	1876	0.378	ug/l #	44
48) Methyl methacrylate	9.26	41	146441	29.293	ug/l #	47
51) 4-Methyl-2-Pentanone	9.98	43	6558	1.125	ug/l	89
54) cis-1,3-Dichloropropene	9.90	75	34976	5.017	ug/l #	69
56) Ethyl methacrylate	10.54	69	1489	0.217	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	6243	0.795	ug/l #	43
59) 2-Hexanone	10.76	43	93747	21.446	ug/l #	54
74) N-amyl acetate	12.04	43	2304	0.297	ug/l #	49
76) 1,2,3-Trichloropropane	12.40	75	161499	31.094	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	161499	109.422	ug/l #	6
95) Naphthalene	15.13	128	1152	Below Cal	#	72
96) 1,2,3-Trichlorobenzene	15.30	180	117	1.023	ug/l #	17


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057870.D
Acq On : 10 Sep 2019 18:08
Operator : JC/SP
Sample : K4623-25
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-C2-(4)

Quant Time: Sep 11 03:55:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.00 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampled :


PAR-11-C2-(4)

Tgt Ion:168 Resp: 290677

Ion Ratio Lower Upper

168 100

99 61.9 48.0 72.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



7.65

100000

50000

0

Time-->



#5

Bromomethane

Concen: Below Cal

RT: 2.55 min Scan# 240

Delta R.T. 0.02 min

Lab File: VN057870.D


Acq: 10 Sep 2019 18:08

Tgt Ion: 94 Resp: 111

Ion Ratio Lower Upper

94 100

96 0.0 77.0 115.4#



Abundance Ion 93.90 (93.60 to 94.60): VN

Ion 95.90 (95.60 to 96.60): VN

2.55

250

200


150


100

50

0

Time-->









#11

Tert butyl alcohol
 Concen: 0.689 ug/l
 RT: 4.72 min Scan# 916
 Delta R.T. -0.05 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-C2-(4)




Tgt Ion: 59 Resp: 491
 Ion Ratio Lower Upper
 59 100
 57 0.0 8.4 12.6#




#16

Acetone
 Concen: 5.452 ug/l
 RT: 3.80 min Scan# 631
 Delta R.T. 0.01 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Tgt Ion: 43 Resp: 10523
 Ion Ratio Lower Upper
 43 100
 58 28.0 22.2 33.2





#17

Carbon Disulfide

Concen: 0.249 ug/l

RT: 4.03 min Scan# 701

Delta R.T. 0.00 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :


PAR-11-C2-(4)

Tgt Ion: 76 Resp: 435

Ion Ratio Lower Upper

76 100

78 0.0 7.1 10.7#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0

4.03

300



200

100

0

4.00 4.02 4.04 4.06

Time-->



#18

Methyl Acetate

Concen: 1.403 ug/l

RT: 4.32 min Scan# 792

Delta R.T. 0.02 min

Lab File: VN057870.D


Acq: 10 Sep 2019 18:08

Tgt Ion: 43 Resp: 7219

Ion Ratio Lower Upper

43 100

74 6.4 17.9 26.9#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0

4.32

2500

2000

1500


1000


500

0

4.25 4.30 4.35 4.40

Time-->









#20
Methylene Chloride
Concen: 0.565 ug/l
RT: 4.53 min Scan# 858
Delta R.T. 0.01 min
Lab File: VN057870.D
Acq: 10 Sep 2019 18:08




Instrument :
MSVOA_N
ClientSampleId :
PAR-11-C2-(4)


Tgt Ion: 84 Resp: 6546
Ion Ratio Lower Upper
84 100
49 108.6 104.8 157.2
51 39.7 33.0 49.4
86 71.1 50.5 75.7



#29
Tetrahydrofuran
Concen: 0.202 ug/l
RT: 7.20 min Scan# 1687
Delta R.T. 0.00 min
Lab File: VN057870.D
Acq: 10 Sep 2019 18:08

Tgt Ion: 42 Resp: 348
Ion Ratio Lower Upper
42 100
72 8.6 33.7 50.5#
71 0.0 31.9 47.9#









#31
Cyclohexane
Concen: Below Cal
RT: 7.65 min Scan# 1827
Delta R.T. 0.01 min
Lab File: VN057870.D
Acq: 10 Sep 2019 18:08

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-C2-(4)

Tgt Ion: 56 Resp: 7757
Ion Ratio Lower Upper
56 100
69 123.2 25.9 38.9#
84 96.5 65.6 98.4




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 55.817 ug/l
RT: 8.01 min Scan# 1941
Delta R.T. 0.00 min
Lab File: VN057870.D
Acq: 10 Sep 2019 18:08

Tgt Ion: 65 Resp: 328619
Ion Ratio Lower Upper
65 100
67 50.7 0.0 104.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.58 min Scan# 2116
 Delta R.T. 0.00 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-C2-(4)

Tgt Ion:114 Resp: 489076
 Ion Ratio Lower Upper
 114 100
 63 22.5 0.0 44.0
 88 15.6 0.0 31.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 48.633 ug/l
 RT: 7.58 min Scan# 1805
 Delta R.T. 0.00 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Tgt Ion:113 Resp: 219527
 Ion Ratio Lower Upper
 113 100
 111 103.5 81.3 121.9
 192 18.1 14.6 22.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.494 ug/l
 RT: 7.65 min Scan# 1828
 Delta R.T. -0.13 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Instrument :
 MSVOA_N
 ClientSampled :
 PAR-11-C2-(4)

Tgt Ion: 75 Resp: 25490
 Ion Ratio Lower Upper
 75 100
 110 8.1 16.3 48.9#
 77 0.0 25.0 37.4#




Abundance Ion 74.90 (74.60 to 75.60): VN0
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VN0






#38
 Carbon Tetrachloride
 Concen: 5.577 ug/l
 RT: 7.65 min Scan# 1828
 Delta R.T. -0.11 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Tgt Ion: 117 Resp: 28882
 Ion Ratio Lower Upper
 117 100
 119 4.9 76.9 115.3#
 121 0.0 23.4 35.2#



Abundance Ion 116.80 (116.50 to 117.50): V
 Ion 118.80 (118.50 to 119.50): V
 Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 9.273 ug/l

RT: 8.16 min Scan# 1985

Delta R.T. 0.00 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

PAR-11-C2-(4)


Tgt Ion: 43 Resp: 91382

Ion Ratio Lower Upper

43 100

61 17.3 20.1 30.1#




87 10.9 9.5 14.3



Abundance Ion 43.00 (42.70 to 43.70): VN057870.D (-1984) (-)

Ion 61.00 (60.70 to 61.70): VN057870.D (-1984) (-)

Ion 87.00 (86.70 to 87.70): VN057870.D (-1984) (-)



#45

1,2-Dichloropropane

Concen: 0.378 ug/l

RT: 9.26 min Scan# 2329

Delta R.T. 0.15 min

Lab File: VN057870.D


Acq: 10 Sep 2019 18:08

Tgt Ion: 63 Resp: 1876

Ion Ratio Lower Upper



63 100


65 0.0 24.8 37.2#



Abundance Ion 62.90 (62.60 to 63.60): VN057870.D (-2329) (-)

Ion 64.90 (64.60 to 65.60): VN057870.D (-2329) (-)





#48

Methyl methacrylate

Concen: 29.293 ug/l

RT: 9.26 min Scan# 2329

Delta R.T. 0.07 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

PAR-11-C2-(4)


Tgt Ion: 41 Resp: 146441

Ion Ratio Lower Upper

41 100

69 0.1 62.7 94.1#




39 59.1 52.7 79.1



Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 39.00 (38.70 to 39.70): VN0



#50

Toluene-d8

Concen: 51.655 ug/l

RT: 10.09 min Scan# 2586

Delta R.T. 0.00 min

Lab File: VN057870.D


Acq: 10 Sep 2019 18:08

Tgt Ion: 98 Resp: 905286

Ion Ratio Lower Upper



98 100


100 63.0 50.3 75.5



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 1.125 ug/l

RT: 9.98 min Scan# 2552

Delta R.T. 0.00 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampled :


PAR-11-C2-(4)

Tgt Ion: 43 Resp: 6558

Ion Ratio Lower Upper




43 100

58 30.8 29.8 44.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#54

cis-1,3-Dichloropropene

Concen: 5.017 ug/l

RT: 9.90 min Scan# 2528

Delta R.T. 0.07 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08


Tgt Ion: 75 Resp: 34976

Ion Ratio Lower Upper

75 100

77 0.0 24.8 37.2#



39 44.7 46.2 69.4#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0





#56

Ethyl methacrylate

Concen: 0.217 ug/l

RT: 10.54 min Scan# 2727

Delta R.T. 0.12 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

PAR-11-C2-(4)


Tgt Ion: 69 Resp: 1489

Ion Ratio Lower Upper

69 100

41 6102.7 56.7 85.1#




39 2854.1 36.3 54.5#



Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 0.795 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.06 min

Lab File: VN057870.D


Acq: 10 Sep 2019 18:08

Tgt Ion: 76 Resp: 6243

Ion Ratio Lower Upper



76 100


78 0.0 25.4 38.2#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#59
 2-Hexanone
 Concen: 21.446 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.01 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08




Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-C2-(4)


Tgt Ion: 43 Resp: 93747
 Ion Ratio Lower Upper
 43 100
 58 18.9 25.6 76.8#



#62
 4-Bromofluorobenzene
 Concen: 43.703 ug/l
 RT: 12.40 min Scan# 3306
 Delta R.T. 0.00 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Tgt Ion: 95 Resp: 291883
 Ion Ratio Lower Upper
 95 100
 174 74.9 0.0 150.2
 176 72.4 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN057870.D
Acq: 10 Sep 2019 18:08

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-C2-(4)

Tgt Ion:117 Resp: 450593
Ion Ratio Lower Upper
117 100
82 59.9 47.2 70.8
119 32.2 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3600
Delta R.T. 0.00 min
Lab File: VN057870.D
Acq: 10 Sep 2019 18:08

Tgt Ion:152 Resp: 229594
Ion Ratio Lower Upper
152 100
115 60.2 30.3 91.0
150 159.4 0.0 352.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-ethyl acetate

Concen: 0.297 ug/l

RT: 12.04 min Scan# 3193

Delta R.T. -0.03 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

PAR-11-C2-(4)

Tgt Ion: 43 Resp: 2304





Ion Ratio Lower Upper

43 100

70 0.0 30.6 45.8#

55 5.8 18.6 27.8#

61 0.0 17.4 26.2#



#76

1,2,3-Trichloropropane

Concen: 31.094 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN057870.D




Acq: 10 Sep 2019 18:08


Tgt Ion: 75 Resp: 161499

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#





#81

trans-1,4-Dichloro-2-butene

Concen: 109.422 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08

Instrument :

MSVOA_N

ClientSampled :

PAR-11-C2-(4)


Tgt Ion: 75 Resp: 161499

Ion Ratio Lower Upper

75 100

53 0.0 88.4 132.6#




89 0.0 35.3 52.9#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#95

Naphthalene

Concen: Below Cal

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN057870.D

Acq: 10 Sep 2019 18:08


Tgt Ion: 128 Resp: 1152

Ion Ratio Lower Upper

128 100

127 0.0 10.3 15.5#



129 2.6 8.6 12.8#




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 1.023 ug/l
 RT: 15.30 min Scan# 4208
 Delta R.T. 0.00 min
 Lab File: VN057870.D
 Acq: 10 Sep 2019 18:08

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-C2-(4)

Tgt Ion	Ratio	Lower	Upper
180	100		
182	186.3	48.3	144.8#
145	0.0	16.3	48.8#

