

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057878.D
 Acq On : 10 Sep 2019 21:05
 Operator : JC/SP
 Sample : K4623-34
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-D2-(4)

Quant Time: Sep 11 04:42:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	292157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	493740	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	455635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	234187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	312208	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
35) Dibromofluoromethane	7.58	113	211252	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
50) Toluene-d8	10.09	98	869128	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
62) 4-Bromofluorobenzene	12.40	95	281104	41.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.38%	
Target Compounds						
16) Acetone	3.80	43	12826	6.612	ug/l	97
43) Isopropyl Acetate	8.16	43	67153	6.750	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057878.D
Acq On : 10 Sep 2019 21:05
Operator : JC/SP
Sample : K4623-34
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-D2-(4)

Quant Time: Sep 11 04:42:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

