

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057879.D
 Acq On : 10 Sep 2019 21:27
 Operator : JC/SP
 Sample : K4623-35
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-D3-(4)

Quant Time: Sep 11 04:44:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	292435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	499801	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	466873	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	237991	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	333763	56.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.70%	
35) Dibromofluoromethane	7.58	113	225614	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	10.09	98	929359	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	12.40	95	299844	43.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.86%	
Target Compounds						
16) Acetone	3.81	43	15108	7.781	ug/l	92
43) Isopropyl Acetate	8.16	43	63391	6.294	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057879.D
Acq On : 10 Sep 2019 21:27
Operator : JC/SP
Sample : K4623-35
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-D3-(4)

Quant Time: Sep 11 04:44:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

