

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057882.D
 Acq On : 10 Sep 2019 22:33
 Operator : JC/SP
 Sample : K4623-39
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-E2-(0)

Quant Time: Sep 11 04:50:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	297988	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	501493	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	463575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	237622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	327517	54.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.54%	
35) Dibromofluoromethane	7.58	113	221091	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
50) Toluene-d8	10.09	98	914395	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	12.40	95	293388	42.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.68%	
Target Compounds						
16) Acetone	3.80	43	11943	6.036	ug/l	99
43) Isopropyl Acetate	8.15	43	70557	6.982	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057882.D
Acq On : 10 Sep 2019 22:33
Operator : JC/SP
Sample : K4623-39
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-E2-(0)

Quant Time: Sep 11 04:50:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

