

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057883.D
 Acq On : 10 Sep 2019 22:55
 Operator : JC/SP
 Sample : K4623-40
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-E3-(0)

Quant Time: Sep 11 04:52:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	293556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	487548	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	449435	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	232105	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	323815	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
35) Dibromofluoromethane	7.58	113	217204	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	10.09	98	891244	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	12.40	95	290347	43.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.22%	
Target Compounds						
16) Acetone	3.80	43	11693	5.999	ug/l	90
43) Isopropyl Acetate	8.16	43	48915	4.979	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057883.D
Acq On : 10 Sep 2019 22:55
Operator : JC/SP
Sample : K4623-40
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-E3-(0)

Quant Time: Sep 11 04:52:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

