

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057884.D
 Acq On : 10 Sep 2019 23:17
 Operator : JC/SP
 Sample : K4624-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-E4-(0)

Quant Time: Sep 11 04:54:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	285024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	498234	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	462950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	236864	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	335190	58.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.12%	
35) Dibromofluoromethane	7.57	113	220293	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
50) Toluene-d8	10.09	98	927542	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	12.41	95	300397	44.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.30%	
Target Compounds						
16) Acetone	3.81	43	19043	10.062	ug/l	99
43) Isopropyl Acetate	8.16	43	49203	4.901	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057884.D
Acq On : 10 Sep 2019 23:17
Operator : JC/SP
Sample : K4624-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-E4-(0)

Quant Time: Sep 11 04:54:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

