

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
 Data File : VN057888.D
 Acq On : 11 Sep 2019 00:45
 Operator : JC/SP
 Sample : K4818-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-MW-16-10.5-091019

Quant Time: Sep 11 05:08:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	267345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	459463	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	413842	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	217903	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	309176	57.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.20%	
35) Dibromofluoromethane	7.58	113	205919	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
50) Toluene-d8	10.09	98	817706	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	12.41	95	270572	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
Target Compounds						
16) Acetone	3.80	43	39101	22.028	ug/l	99
18) Methyl Acetate	4.31	43	4592	0.970	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091119\
Data File : VN057888.D
Acq On : 11 Sep 2019 00:45
Operator : JC/SP
Sample : K4818-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-MW-16-10.5-091019

Quant Time: Sep 11 05:08:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

