

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057900.D
 Acq On : 11 Sep 2019 5:54
 Operator : JC/SP
 Sample : K4820-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
Client Sampled :
 GW200A-B4-47.0-52.0-091019

Manual Integrations
APPROVED
 MMDadoda
 9/13/2019 7:07:02 AM

Quant Time: Sep 11 07:28:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	284710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	487827	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	439931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	222077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	324438	56.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.52%	
35) Dibromofluoromethane	7.58	113	224037	49.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.52%	
50) Toluene-d8	10.09	98	899827	51.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.94%	
62) 4-Bromofluorobenzene	12.40	95	292185	43.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.80	96	4473m	1.018	ug/l	Qvalue
44) Trichloroethene	8.83	130	38610	9.242	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057900.D
 Acq On : 11 Sep 2019 5:54
 Operator : JC/SP
 Sample : K4820-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
Client Sampled :
 GW200A-B4-47.0-52.0-091019

Manual Integrations
APPROVED
 MMDadoda
 9/13/2019 7:07:02 AM

Quant Time: Sep 11 07:28:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

