

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057904.D
 Acq On : 11 Sep 2019 7:22
 Operator : JC/SP
 Sample : K4624-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-E2-(4)

Quant Time: Sep 11 08:04:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	293775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	510202	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	465279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	236819	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	327808	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
35) Dibromofluoromethane	7.58	113	223263	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
50) Toluene-d8	10.09	98	901357	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	12.40	95	293661	42.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.30%	
Target Compounds						
16) Acetone	3.81	43	10816	5.545	ug/l	94
43) Isopropyl Acetate	8.15	43	51004	4.961	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091119\
Data File : VN057904.D
Acq On : 11 Sep 2019 7:22
Operator : JC/SP
Sample : K4624-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-E2-(4)

Quant Time: Sep 11 08:04:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

