

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057908.D
 Acq On : 11 Sep 2019 8:49
 Operator : JC/SP
 Sample : K4664-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-090619

Quant Time: Sep 13 05:44:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	297934	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	509781	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	474714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	247869	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	332487	55.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.20%	
35) Dibromofluoromethane	7.58	113	228427	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	10.09	98	947670	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.40	95	310213	44.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.12%	
Target Compounds						
16) Acetone	3.81	43	8603	4.349	ug/l	93
20) Methylene Chloride	4.53	84	23675	5.183	ug/l	97
43) Isopropyl Acetate	8.16	43	80234	7.811	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091119\
Data File : VN057908.D
Acq On : 11 Sep 2019 8:49
Operator : JC/SP
Sample : K4664-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-090619

Quant Time: Sep 13 05:44:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

