

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057911.D
 Acq On : 11 Sep 2019 9:56
 Operator : JC/SP
 Sample : K4643-31
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-C1-(0)

Quant Time: Sep 12 01:44:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	319459	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	534326	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	487170	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	254191	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	337149	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
35) Dibromofluoromethane	7.58	113	230271	46.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.38%	
50) Toluene-d8	10.09	98	948824	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	12.40	95	311243	42.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.30%	
Target Compounds						
16) Acetone	3.80	43	11029	5.200	ug/l	98
43) Isopropyl Acetate	8.16	43	39624	3.680	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091119\
Data File : VN057911.D
Acq On : 11 Sep 2019 9:56
Operator : JC/SP
Sample : K4643-31
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-C1-(0)

Quant Time: Sep 12 01:44:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

