

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057917.D
 Acq On : 11 Sep 2019 12:07
 Operator : JC/SP
 Sample : K4643-40
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-D1-(0)

Quant Time: Sep 12 01:57:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	285140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	494658	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	443859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	227963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	330100	57.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.32%	
35) Dibromofluoromethane	7.58	113	224924	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
50) Toluene-d8	10.09	98	890874	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
62) 4-Bromofluorobenzene	12.40	95	292744	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
Target Compounds						
16) Acetone	3.80	43	10714	5.659	ug/l	92
20) Methylene Chloride	4.53	84	9244	1.367	ug/l	89
43) Isopropyl Acetate	8.16	43	13038	1.308	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091119\
Data File : VN057917.D
Acq On : 11 Sep 2019 12:07
Operator : JC/SP
Sample : K4643-40
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 60 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-D1-(0)

Quant Time: Sep 12 01:57:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

