

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057918.D
 Acq On : 11 Sep 2019 12:29
 Operator : JC/SP
 Sample : K4644-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-D3-(0)

Quant Time: Sep 12 01:58:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	290836	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	493648	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	452020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	229805	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	330416	56.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.18%	
35) Dibromofluoromethane	7.58	113	225902	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	10.09	98	892884	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	12.40	95	293145	43.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.98%	
Target Compounds						
16) Acetone	3.81	43	10698	5.540	ug/l #	84
43) Isopropyl Acetate	8.16	43	16910	1.700	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091119\
Data File : VN057918.D
Acq On : 11 Sep 2019 12:29
Operator : JC/SP
Sample : K4644-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 61 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-D3-(0)

Quant Time: Sep 12 01:58:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

