

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091119\
 Data File : VN057921.D
 Acq On : 11 Sep 2019 13:35
 Operator : JC/SP
 Sample : K4622-38
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAR-11-B4-(4)

Quant Time: Sep 12 02:03:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	281771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	485360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	442914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	317420	55.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.24%	
35) Dibromofluoromethane	7.58	113	217000	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
50) Toluene-d8	10.09	98	851458	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	12.41	95	284879	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
Target Compounds						
16) Acetone	3.80	43	14293	7.640	ug/l	95
20) Methylene Chloride	4.53	84	15585	3.224	ug/l	95
43) Isopropyl Acetate	8.15	43	84319	8.621	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091119\
Data File : VN057921.D
Acq On : 11 Sep 2019 13:35
Operator : JC/SP
Sample : K4622-38
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAR-11-B4-(4)

Quant Time: Sep 12 02:03:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

