

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091120\
 Data File : VN063321.D
 Acq On : 11 Sep 2020 17:54
 Operator : JC/MD
 Sample : L3988-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-PZ193I-20200909

Quant Time: Sep 12 04:42:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 11 06:02:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	364326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	685853	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	668825	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	269301	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	289102	50.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.02%	
35) Dibromofluoromethane	7.55	113	222974	50.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.78%	
50) Toluene-d8	10.06	98	873484	51.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.02%	
62) 4-Bromofluorobenzene	12.38	95	331690	52.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.92%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091120\
Data File : VN063321.D
Acq On : 11 Sep 2020 17:54
Operator : JC/MD
Sample : L3988-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-PZ193I-20200909

Quant Time: Sep 12 04:42:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 11 06:02:56 2020
Response via : Initial Calibration

