


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091120\
 Data File : VN063304.D
 Acq On : 11 Sep 2020 10:03
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
 Title : SW846 8260
 Last Update : Fri Sep 11 06:02:56 2020



AutoFind: Scans 3308, 3309, 3310; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	20968	PASS
75	95	30	60	50.7	51136	PASS
95	95	100	100	100.0	100856	PASS
96	95	5	9	6.6	6663	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.6	72197	PASS
175	174	5	9	8.2	5896	PASS
176	174	95	101	96.5	69672	PASS
177	176	5	9	6.2	4293	PASS