

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091123\
 Data File : VN079228.D
 Acq On : 11 Sep 2023 19:54
 Operator : JC\MD
 Sample : 04299-07ME
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-04-7.0-7.5ME

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/12/2023
 Supervised By : Mahesh Dadoda 09/12/2023

Quant Time: Sep 12 05:22:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	113393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	200846	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.871	117	195393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	79253	51.776	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.560%			
35) Dibromofluoromethane	8.165	113	69343	51.904	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.800%			
50) Toluene-d8	10.571	98	252445	51.461	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.920%			
62) 4-Bromofluorobenzene	12.853	95	91297	58.345	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 116.700%			
Target Compounds						
						Qvalue
67) Ethyl Benzene	11.965	91	63978	8.450	ug/l	98
68) m/p-Xylenes	12.076	106	21116	7.482	ug/l	97
69) o-Xylene	12.400	106	26995	9.974	ug/l	97
73) Isopropylbenzene	12.700	105	10178	1.242	ug/l	98
80) 1,3,5-Trimethylbenzene	13.176	105	43948	6.437	ug/l	99
84) 1,2,4-Trimethylbenzene	13.488	105	52894	8.135	ug/l	99
86) p-Isopropyltoluene	13.729	119	10121m	1.629	ug/l	
95) Naphthalene	15.647	128	1454522	405.663	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091123\
Data File : VN079228.D
Acq On : 11 Sep 2023 19:54
Operator : JC\MD
Sample : 04299-07ME
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-04-7.0-7.5ME

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/12/2023
Supervised By :Mahesh Dadoda 09/12/2023

Quant Time: Sep 12 05:22:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

