

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091123\
 Data File : VN079229.D
 Acq On : 11 Sep 2023 20:18
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 12 05:23:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	125945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	229304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	218434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	87317	51.359	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.720%	
35) Dibromofluoromethane	8.171	113	77402	50.746	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.500%	
50) Toluene-d8	10.570	98	285783	51.027	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.060%	
62) 4-Bromofluorobenzene	12.853	95	97941	54.823	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.640%	
Target Compounds						
95) Naphthalene	15.647	128	24965	7.392	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091123\
Data File : VN079229.D
Acq On : 11 Sep 2023 20:18
Operator : JC\MD
Sample : IBLK
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 12 05:23:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

