

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091124\
 Data File : VN083814.D
 Acq On : 11 Sep 2024 22:45
 Operator : JC\MD
 Sample : P3913-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 9424-D-F

Quant Time: Sep 12 02:17:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	116523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	252165	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	151200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	95645	54.436	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.880%	
35) Dibromofluoromethane	8.165	113	75713	46.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.800%	
50) Toluene-d8	10.565	98	284253	50.934	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.860%	
62) 4-Bromofluorobenzene	12.847	95	141878	61.619	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 123.240%#	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.541	59	7473	31.762	ug/l #	87
16) Acetone	4.436	43	154345	219.134	ug/l	98
17) Carbon Disulfide	4.694	76	10289	2.918	ug/l #	91
18) Methyl Acetate	5.030	43	4682	1.386	ug/l #	78
25) 2-Butanone	7.482	43	144741	156.351	ug/l	100
52) Toluene	10.635	92	4942	1.105	ug/l	93
95) Naphthalene	15.635	128	67766	8.093	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091124\
Data File : VN083814.D
Acq On : 11 Sep 2024 22:45
Operator : JC\MD
Sample : P3913-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
9424-D-F

Quant Time: Sep 12 02:17:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

