

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091218\
 Data File : VN051151.D
 Acq On : 12 Sep 2018 14:35
 Operator : MD\SY
 Sample : J4856-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-3

Quant Time: Sep 13 08:24:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	598557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	972367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	831743	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	306240	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	378489	56.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.64%	
35) Dibromofluoromethane	7.59	113	361648	49.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.00%	
50) Toluene-d8	10.09	98	1312817	47.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.70%	
62) 4-Bromofluorobenzene	12.40	95	368506	40.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.36%	
Target Compounds						
16) Acetone	3.83	43	32676	14.34	ug/l	96
18) Methyl Acetate	4.34	43	12865	2.41	ug/l	93
20) Methylene Chloride	4.55	84	97691	13.28	ug/l	97
43) Isopropyl Acetate	8.17	43	297404	27.26	ug/l #	83
95) Naphthalene	15.13	128	62476	10.19	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091218\
Data File : VN051151.D
Acq On : 12 Sep 2018 14:35
Operator : MD\SY
Sample : J4856-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1-3

Quant Time: Sep 13 08:24:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

