


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091218\
Data File : VN051136.D
Acq On : 12 Sep 2018 7:22
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Title : SW846 8260
Last Update : Tue Sep 11 02:27:09 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	4966	PASS
75	95	30	60	42.4	13554	PASS
95	95	100	100	100.0	31960	PASS
96	95	5	9	7.0	2230	PASS
173	174	0.00	2	1.9	551	PASS
174	95	50	100	92.1	29429	PASS
175	174	5	9	8.1	2382	PASS
176	174	95	101	96.0	28258	PASS
177	176	5	9	6.3	1775	PASS