

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091219\
 Data File : VN057930.D
 Acq On : 11 Sep 2019 17:58
 Operator : JC/SP
 Sample : PB122748TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB122748TB

Quant Time: Sep 12 03:18:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	311484	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	521352	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	485272	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	246891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	345494	54.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.52%	
35) Dibromofluoromethane	7.58	113	229783	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	10.09	98	947688	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
62) 4-Bromofluorobenzene	12.41	95	311501	43.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.50%	
Target Compounds						
43) Isopropyl Acetate	8.16	43	95941	9.133	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091219\
Data File : VN057930.D
Acq On : 11 Sep 2019 17:58
Operator : JC/SP
Sample : PB122748TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB122748TB

Quant Time: Sep 12 03:18:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

