

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091219\
 Data File : VN057938.D
 Acq On : 11 Sep 2019 20:54
 Operator : JC/SP
 Sample : K4818-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-091019-01

Quant Time: Sep 12 03:35:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	276922	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	461030	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	421663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	215679	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	298892	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
35) Dibromofluoromethane	7.58	113	202251	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
50) Toluene-d8	10.09	98	819079	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.41	95	263600	41.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.74%	
Target Compounds						
16) Acetone	3.81	43	27236	14.813	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091219\
Data File : VN057938.D
Acq On : 11 Sep 2019 20:54
Operator : JC/SP
Sample : K4818-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-091019-01

Quant Time: Sep 12 03:35:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

