

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091219\
 Data File : VN057939.D
 Acq On : 11 Sep 2019 21:16
 Operator : JC/SP
 Sample : K4836-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP

Quant Time: Sep 12 03:38:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	275120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	456045	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	414987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	208222	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	298702	53.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.20%	
35) Dibromofluoromethane	7.58	113	205695	48.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.74%	
50) Toluene-d8	10.09	98	808130	49.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.41	95	265078	42.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091219\
Data File : VN057939.D
Acq On : 11 Sep 2019 21:16
Operator : JC/SP
Sample : K4836-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP

Quant Time: Sep 12 03:38:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

