

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091219\
 Data File : VN057941.D
 Acq On : 11 Sep 2019 22:01
 Operator : JC/SP
 Sample : K4835-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PZ-3

Quant Time: Sep 12 03:42:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	277732	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	377308	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	341404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	206252	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	314013	55.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.64%	
35) Dibromofluoromethane	7.58	113	213161	61.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.42%	
50) Toluene-d8	10.09	98	6042	0.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.90%	
62) 4-Bromofluorobenzene	12.41	95	208682	40.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.00%	
Target Compounds						
11) Tert butyl alcohol	4.77	59	499453	733.590	ug/l	100
16) Acetone	3.80	43	125104	67.842	ug/l	98
19) Methyl tert-butyl Ether	5.02	73	1458134	118.476	ug/l	98
25) 2-Butanone	6.83	43	13723	5.142	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091219\
Data File : VN057941.D
Acq On : 11 Sep 2019 22:01
Operator : JC/SP
Sample : K4835-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PZ-3

Quant Time: Sep 12 03:42:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

