

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091219\
 Data File : VN057947.D
 Acq On : 12 Sep 2019 00:13
 Operator : JC/SP
 Sample : K4848-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CBH-139-A

Quant Time: Sep 12 03:54:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	279219	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	473051	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	421881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	215951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	317157	56.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.16%	
35) Dibromofluoromethane	7.58	113	214795	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
50) Toluene-d8	10.09	98	859148	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	12.41	95	279519	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	
Target Compounds						
16) Acetone	3.81	43	14137	7.625	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091219\
Data File : VN057947.D
Acq On : 12 Sep 2019 00:13
Operator : JC/SP
Sample : K4848-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CBH-139-A

Quant Time: Sep 12 03:54:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

