

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
 Data File : VN074412.D
 Acq On : 12 Sep 2022 14:34
 Operator : JC\MD
 Sample : N4471-03DL 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11A-11-16-083022DL

Quant Time: Sep 13 05:55:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	296340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	515058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	498916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	225867	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	229856	44.831	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.660%
35) Dibromofluoromethane	7.951	113	185984	53.488	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.980%
50) Toluene-d8	10.380	98	620166	45.774	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.540%
62) 4-Bromofluorobenzene	12.674	95	243210	49.443	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.880%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.134	101	5721	1.507	ug/l #	73
27) cis-1,2-Dichloroethene	7.245	96	114411	22.434	ug/l	89
44) Trichloroethene	9.151	130	107411	27.317	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
Data File : VN074412.D
Acq On : 12 Sep 2022 14:34
Operator : JC\MD
Sample : N4471-03DL 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11A-11-16-083022DL

Quant Time: Sep 13 05:55:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

