

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
 Data File : VN074413.D
 Acq On : 12 Sep 2022 15:54
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 13 05:55:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.998	168	293975	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.892	114	518344	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.686	117	549510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	293707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	220177	43.289	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.580%
35) Dibromofluoromethane	7.945	113	183450	52.424	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.840%
50) Toluene-d8	10.375	98	603813	44.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.560%
62) 4-Bromofluorobenzene	12.680	95	274545	55.459	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.920%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.139	130	17376	4.391	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
Data File : VN074413.D
Acq On : 12 Sep 2022 15:54
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 13 05:55:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

