

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
 Data File : VN074419.D
 Acq On : 12 Sep 2022 19:36
 Operator : JC\MD
 Sample : N4558-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-EB-20220908-

Quant Time: Sep 13 06:28:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	289295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	502789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	479099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	211360	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	221545	44.263	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.520%
35) Dibromofluoromethane	7.951	113	182529	53.775	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.540%
50) Toluene-d8	10.380	98	605153	45.755	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.520%
62) 4-Bromofluorobenzene	12.674	95	233193	48.563	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.120%
Target Compounds						
					Qvalue	
30) Chloroform	7.739	83	10711	1.311	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
Data File : VN074419.D
Acq On : 12 Sep 2022 19:36
Operator : JC\MD
Sample : N4558-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-EB-20220908-

Quant Time: Sep 13 06:28:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

